
CERTIFICATE OF ANALYSIS

HEMOGLOBIN A1C CONTROL (NORMAL)

COD 18001 **LOT** 0006183 **EXP** 2028-10-07 **MANUF** 2026-06-25

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
N	0005A	2028-10-07

CONTROL DATA

CONTROL MATERIAL

ASSIGNED VALUES

Component	Unit	Value	Range (min)	Range (max)
HEMOGLOBIN A1C	mmol/mol (IFCC)	38	30	46
HEMOGLOBIN A1C	% (NGSP)	5,7	4,6	6,8

Note: Components from human origin have been tested and found to be negative for the presence of antibodies anti-HIV (I and II) and anti-HCV as well as for HBs antigen. However the product should be handled cautiously as potentially infectious.

This lot meets the quality requirements, and it is released for sale

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/06/30

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes

--

