

Certificate of Analysis

PROTEIN (TOTAL)

COD	21513	LOT	0006110	EXP	2028-04-25	MANUF	2026-06-15
-----	-------	-----	---------	-----	------------	-------	------------

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00006	2028-04-25
B	00006	2028-04-25

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,040	0,000	0,150	Pass
ACCURACY SAMPLE 1	0004B	g/L	51,310	42,100	57,300	Pass
ACCURACY SAMPLE 2	0003A	g/L	82,861	72,500	92,300	Pass
ACCURACY SAMPLE 3	0001A	g/L	57,147	49,600	63,200	Pass
ACCURACY SAMPLE 4	0002A	g/L	92,325	80,700	102,700	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/06/17

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes