

Certificate of Analysis

TRIGLYCERIDES

COD

21528

LOT

0006138

EXP

2027-11-12

MANUF

2026-06-18

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00012	2027-11-12

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,011	0,000	0,150	Pass
ACCURACY SAMPLE 1	0003A	mg/dL	225,071	188,000	254,000	Pass
ACCURACY SAMPLE 2	0002A	mg/dL	129,284	110,000	148,000	Pass
ACCURACY SAMPLE 3	0001A	mg/dL	131,595	107,000	145,000	Pass
ACCURACY SAMPLE 4	045XA	mg/dL	194,073	166,000	224,000	Pass

Spec.:Specification

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/06/22

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes