

Certificate of Analysis

LACTATE DEHYDROGENASE (LDH)

COD

21580

LOT

0005719

EXP

2028-11-04

MANUF

2026-05-07

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00004	2028-11-04
B	00004	2028-11-04

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	1,934	1,200		Pass
ACCURACY SAMPLE 1	0003A	U/L	415,244	344,000	494,000	Pass
ACCURACY SAMPLE 2	0001A	U/L	819,908	672,000	968,000	Pass
ACCURACY SAMPLE 3	0001A	U/L	495,729	416,000	598,000	Pass
ACCURACY SAMPLE 4	0001A	U/L	829,762	693,000	997,000	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/05/12

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes