

Certificate of Analysis

CHOLESTEROL HDL DIRECT

COD

21594

LOT

0006213

EXP

2027-11-30

MANUF

2026-06-29

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00013	2027-11-30
B	00013	2027-11-30

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,005	0,000	0,050	Pass
ACCURACY SAMPLE 1	002V	mg/dL	14,593	-10,000	50,000	Pass
ACCURACY SAMPLE 2	0002A	mg/dL	47,039	35,500	59,100	Pass
ACCURACY SAMPLE 3	0002B	mg/dL	40,609	31,100	51,800	Pass
ACCURACY SAMPLE 4	0002A	mg/dL	63,127	46,800	78,000	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/07/02

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes