

Certificate of Analysis

LIPASE

COD

21760

LOT

0006256

EXP

2027-10-27

MANUF

2026-07-03

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00028	2027-10-27
B	00028	2027-10-27

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,077	0,000	0,800	Pass
ACCURACY SAMPLE 1	0003A	U/L	92,892	76,000	126,000	Pass
ACCURACY SAMPLE 2	0001A	U/L	49,400	35,100	58,500	Pass
ACCURACY SAMPLE 3	0002A	U/L	79,082	59,700	99,500	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/07/07

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes