

Certificate of Analysis

HEMOGLOBIN A1C-DIRECT (HbA1C-DIR)

COD

22147

LOT

0006265

EXP

2027-12-10

MANUF

2026-07-03

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00011	2027-12-10
B	00011	2027-12-10

CONTROL DATA

CALIBRATION CURVE

Calibrator	Lot	Concentration	Unit	Absorbance
HbA1C-DIR STD Level 1	0006A	33	mmol/mol	0,031
HbA1C-DIR STD Level 2	0006A	67	mmol/mol	0,106
HbA1C-DIR STD Level 3	0006A	95	mmol/mol	0,196
HbA1C-DIR STD Level 4	0006A	115	mmol/mol	0,255

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,385	0,000	0,700	Pass
ACCURACY SAMPLE 1	0001A	mmol/mol	37,119	30,000	44,000	Pass
ACCURACY SAMPLE 2	0002A	mmol/mol	98,119	78,000	116,000	Pass

Spec.:Specification

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/07/08

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes