

Certificate of Analysis

CK-MB CONTROL SERUM

COD 18024 LOT 0005985 EXP 2029-05-17 MANUF 2026-06-02

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
I	0002A	2029-05-17

CONTROL MATERIAL - ASSIGNED VALUES

Component	Unit 1	Value 1	Range (min)	Range (max)
CK	U/L CK	172	146	198
CK-MB	U/L CK-MB	41.5	31.1	51.9

Note: Components from human origin have been tested and found to be negative for the presence of antibodies anti-HIV (I and II) and anti-HCV as well as for HBs antigen. However the product should be handled cautiously as potentially infectious.

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/06/04

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes