

LIPID CONTROL SERUM

NIVEL: I

LOTE: 0003

ESPAÑOL

COMPONENTE	MÉTODO	VALOR	LIMITES	1S	UNIDAD	TRAZABILIDAD
APOLIPOPROTEÍNA A-I	Turbidimetría	140	98 - 182	14	mg/dL	WHO/IFCC SP01-01
		1,40	0,98 - 1,82	0,14	g/L	
APOLIPOPROTEÍNA B	Turbidimetría	161	121 - 201	13	mg/dL	WHO/IFCC SP03-07
		1,61	1,21 - 2,01	0,13	g/L	
COLESTEROL	Colesterol oxidasa/peroxidasa	256	218 - 294	13	mg/dL	SRM 909 (NIST)
		6,62	5,63 - 7,61	0,33	mmol/L	
COLESTEROL HDL	Directo detergente	54,6	41,0 - 68,3	4,6	mg/dL	CDC Reference Method, BMC
		1,42	1,07 - 1,78	0,12	mmol/L	
	Directo TOOS	46,4	34,8 - 58,0	3,9	mg/dL	CDC Reference Method, BMC
		1,20	0,90 - 1,50	0,10	mmol/L	
COLESTEROL LDL	Directo detergente	124	93 - 155	10	mg/dL	CDC Reference Method, BMC
		3,22	2,42 - 4,03	0,27	mmol/L	
	Directo TOOS	171	128 - 214	14	mg/dL	CDC Reference Method, BMC
		4,43	3,32 - 5,54	0,37	mmol/L	
TRIGLICÉRIDOS	Glicerol fosfato oxidasa/peroxidasa	147	125 - 169	7	mg/dL	SRM 909 (NIST)
		1,66	1,41 - 1,91	0,08	mmol/L	

NOTAS:
BMC: Calibrador maestro de BioSystems.