

LIPID CONTROL SERUM

NIVEL: **II**

LOTE: **0003**

ESPAÑOL

COMPONENTE	MÉTODO	VALOR	LIMITES	1S	UNIDAD	TRAZABILIDAD
APOLIPOPROTEÍNA A-I	Turbidimetría	65,9	46,1 - 85,7	6,6	mg/dL	WHO/IFCC SP01-01
		0,659	0,461 - 0,857	0,066	g/L	
APOLIPOPROTEÍNA B	Turbidimetría	208	156 - 260	17	mg/dL	WHO/IFCC SP03-07
		2,08	1,56 - 2,60	0,17	g/L	
COLESTEROL	Colesterol oxidasa/peroxidasa	280	238 - 322	14	mg/dL	SRM 909 (NIST)
		7,25	6,16 - 8,34	0,36	mmol/L	
COLESTEROL HDL	Directo detergente	30,0	22,5 - 37,5	2,5	mg/dL	CDC Reference Method, BMC
		0,777	0,583 - 0,971	0,065	mmol/L	
	Directo TOOS	20,4	15,3 - 25,5	1,7	mg/dL	CDC Reference Method, BMC
		0,529	0,397 - 0,661	0,044	mmol/L	
COLESTEROL LDL	Directo detergente	145	109 - 181	12	mg/dL	CDC Reference Method, BMC
		3,76	2,82 - 4,70	0,31	mmol/L	
	Directo TOOS	201	151 - 251	17	mg/dL	CDC Reference Method, BMC
		5,22	3,92 - 6,53	0,44	mmol/L	
TRIGLICÉRIDOS	Glicerol fosfato oxidasa/peroxidasa	172	146 - 198	9	mg/dL	SRM 909 (NIST)
		1,95	1,66 - 2,24	0,10	mmol/L	

NOTAS:

BMC: Calibrador maestro de BioSystems.

