

Certificate of Analysis

gamma-GLUTAMYLTRANSFERASE (gamma-GT)

COD

21520

LOT

0005204

EXP

2028-07-24

MANUF

2026-03-11

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00007	2028-07-24
B	00007	2028-07-24

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,687	0,000	1,450	Pass
ACCURACY SAMPLE 1	0002A	U/L	41,561	32,900	47,300	Pass
ACCURACY SAMPLE 2	098XB	U/L	185,638	151,000	217,000	Pass
ACCURACY SAMPLE 3	0001A	U/L	34,365	31,400	45,200	Pass
ACCURACY SAMPLE 4	0001A	U/L	146,299	121,000	175,000	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/03/16

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes