

Certificate of Analysis

CREATINE KINASE-MB (CK-MB)

COD

21792

LOT

0005655

EXP

2028-06-25

MANUF

2026-04-30

PRODUCT COMPOSITION

COMPONENT	LOT	EXP
A	00006	2028-06-25
B	00006	2028-06-25

CONTROL DATA

	Lot	Unit	Result	Spec. (min)	Spec. (max)	Pass/Fail/NA
REAGENT ABSORBANCE		Abs	0,105	0,000	0,400	Pass
ACCURACY SAMPLE 1	007V	U/L	16,887	-10,000	30,000	Pass
ACCURACY SAMPLE 2	091XA	U/L	55,852	42,200	70,300	Pass
ACCURACY SAMPLE 3	019XA	U/L	101,797	76,000	127,000	Pass
Spec.:Specification						

This lot meets the quality requirements, and it is released for sale.

Approved by: TECHNICAL RESPONSIBLE / DEPUTY TECHNICAL RESPONSIBLE Date: 2026/05/06

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Notes