

Modbus Map: HydroAct 4 (75FC:4543:F24F:4856)

Input Registers (Read Only)					
Address	Size	Type	Device	Description	Ref.
0	2	FLOAT	UV-254 1.1	Official Value - UVA	
2	2	FLOAT	UV-254 1.1	Official Value - UVT	
4	2	FLOAT	UV-254 1.1	Official Value - Compensated UVA	
6	2	FLOAT	UV-254 1.1	Official Value - COD	
8	2	FLOAT	UV-254 1.1	Official Value - TOC	
10	2	FLOAT	Particle Counter 1.2	Official Value - Cell	
12	2	FLOAT	Particle Counter 1.2	Official Value - Laser	
14	2	FLOAT	Particle Counter 1.2	Official Value - Total Counts	
16	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 1	
18	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 2	
20	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 3	
22	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 4	
24	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 5	
26	2	FLOAT	Particle Counter 1.2	Official Value - Counts Bin 6	
28	2	FLOAT	Particle Counter 1.2	Official Value - Total PPB	
30	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 1	
32	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 2	
34	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 3	
36	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 4	
38	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 5	
40	2	FLOAT	Particle Counter 1.2	Official Value - PPB Bin 6	
42	2	FLOAT	Particle Counter 1.3	Official Value - Cell	
44	2	FLOAT	Particle Counter 1.3	Official Value - Laser	
46	2	FLOAT	Particle Counter 1.3	Official Value - Total Counts	
48	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 1	
50	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 2	
52	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 3	
54	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 4	
56	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 5	
58	2	FLOAT	Particle Counter 1.3	Official Value - Counts Bin 6	
60	2	FLOAT	Particle Counter 1.3	Official Value - Total PPB	
62	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 1	
64	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 2	
66	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 3	
68	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 4	
70	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 5	
72	2	FLOAT	Particle Counter 1.3	Official Value - PPB Bin 6	

Holding Registers (Read Write)					
Address	Size	Type	Device	Description	Ref.
1000	1	U8	HydroAct 4	Alarm state	1

References
<div><div>1. HydroAct 4 - Alarm state</div><div>0 - Normal, 1 - Alarm Active, 2 - Alarm Acknowledged</div></div>