



DIMENSIONAL PART REPORT										LABORATORY				
Report Number:										Result: aceptado				
Customer:										Issue date: December 3, 2023				
Customer part number:														
Part name: ESCALA RECTA_				Evaluation reason: <input type="checkbox"/> Annual validation: <input type="checkbox"/> New part. <input type="checkbox"/> Engineering change. <input type="checkbox"/> New tooling. <input type="checkbox"/> Corrected tooling. <input type="checkbox"/> Prototype piece. <input checked="" type="checkbox"/> Others (Specify): VERIF. ESCALA						Evaluated by: Angelica Vidal				
Part number: ER-057										Evaluation date: December 3, 2023				
Number of pieces: 1										Authorized by: Mireya Hernandez				
Plane No.: _____ Dated: _____ Change: _____ Date of production and/or Batch Number: _____										Evaluation machine: SISTEMA DE VISION (EM-019)				
										Calibration date: June-23				
Observations: Rango 0 a 150 mm / resolución 0.5 mm										PROX. VERIF.=DIC-24				
Item	Measuring equipment used	Characteristics	Especial Features	Spec.:			Limits		Draw: 0			Change: 0		Result
				NOMINAL	TOL. MIN -	TOL. MAX +	MIN. -	MAX. +	Obtained Value:					
									Date 1	Date 2	Date 3	PROM.	DESV.	
1				0.500	0.005	0.005	0.495	0.505	0.496	0.502	0.501	0.500	0.000	OK
2				1.000	0.005	0.005	0.995	1.005	1.003	1.004	1.004	1.004	-0.004	OK
3				1.500	0.005	0.005	1.495	1.505	1.500	1.498	1.503	1.500	0.000	OK
4				2.000	0.005	0.005	1.995	2.005	1.997	1.996	2.002	1.998	0.002	OK
5				2.500	0.005	0.005	2.495	2.505	2.502	2.501	2.502	2.502	-0.002	OK
6				3.000	0.005	0.005	2.995	3.005	3.003	3.002	2.999	3.001	-0.001	OK
7				3.500	0.005	0.005	3.495	3.505	3.504	3.498	3.502	3.501	-0.001	OK
8				4.000	0.005	0.005	3.995	4.005	3.998	3.998	4.002	3.999	0.001	OK
9				4.500	0.005	0.005	4.495	4.505	4.496	4.499	4.498	4.498	0.002	OK
10				5.000	0.005	0.005	4.995	5.005	5.000	5.000	4.999	5.000	0.000	OK
11				5.500	0.005	0.005	5.495	5.505	5.497	5.501	5.500	5.499	0.001	OK
12				6.000	0.005	0.005	5.995	6.005	6.001	5.998	5.997	5.999	0.001	OK
13				6.500	0.005	0.005	6.495	6.505	6.496	6.503	6.497	6.499	0.001	OK
14				7.000	0.005	0.005	6.995	7.005	6.999	6.997	7.001	6.999	0.001	OK
15				7.500	0.005	0.005	7.495	7.505	7.502	7.498	7.496	7.499	0.001	OK
16				8.000	0.005	0.005	7.995	8.005	8.002	7.999	8.000	8.000	0.000	OK
17				8.500	0.005	0.005	8.495	8.505	8.504	8.498	8.498	8.500	0.000	OK
18				9.000	0.005	0.005	8.995	9.005	8.997	9.002	8.997	8.999	0.001	OK
19				9.500	0.005	0.005	9.495	9.505	9.496	9.497	9.497	9.497	0.003	OK
20				10.000	0.005	0.005	9.995	10.005	9.998	10.001	10.003	10.001	-0.001	OK
21				20.000	0.005	0.005	19.995	20.005	19.998	20.001	19.998	19.999	0.001	OK
22				30.000	0.005	0.005	29.995	30.005	30.004	29.998	29.999	30.000	0.000	OK
23				40.000	0.005	0.005	39.995	40.005	40.000	40.004	39.997	40.000	0.000	OK
24				50.000	0.005	0.005	49.995	50.005	49.999	49.996	50.001	49.999	0.001	OK
25				80.000	0.005	0.005	79.995	80.005	80.001	80.003	80.003	80.002	-0.002	OK
26				120.000	0.005	0.005	119.995	120.005	120.000	120.000	119.998	119.999	0.001	OK
25				150.000	0.005	0.005	149.995	150.005	150.003	149.998	150.001	150.001	-0.001	OK

