URINE STRIP IMAGE ANALYSIS REPORT
Generated from data/results/sample_strip2_results.json
Clinical interpretation is not performed.

Sample 2: data/raw_images/sample_strip2.png
Timestamp: 2026-10-31T19:03:09.749322 | Pads detected: 4 | Processing: 0.0355 s
Method: HSV saturation threshold, morphological operations, and contours
  Pad 1 | Parameter: LEUKOCYTES | Colour: Greenish | RGB: 100, 180, 80
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 2 | Parameter: NITRITE | Colour: Yellowish / brownish | RGB: 230, 190, 50
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 3 | Parameter: UROBILINOGEN | Colour: Reddish | RGB: 220, 70, 70
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 4 | Parameter: PROTEIN | Colour: Bluish | RGB: 70, 100, 220
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7

Limitations: Results depend on lighting, camera, and calibration.
Parameter labels follow the configured top-to-bottom order; confirm this mapping for the exact strip model.
Clinical values remain unassigned until calibrated reference data is supplied.
Automated detection has not been validated on real strip photographs.