URINE STRIP IMAGE ANALYSIS REPORT
Generated from data/results/sample_strip3_results.json
Clinical interpretation is not performed.

Sample 3: data/raw_images/sample_strip3.png
Timestamp: 2026-11-01T01:05:48.347458 | Pads detected: 4 | Processing: 0.0176 s
Method: HSV saturation threshold, morphological operations, and contours
  Pad 1 | Parameter: LEUKOCYTES | Colour: Pinkish / purplish | RGB: 180, 70, 180
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 2 | Parameter: NITRITE | Colour: Reddish | RGB: 240, 150, 40
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 3 | Parameter: UROBILINOGEN | Colour: Bluish | RGB: 50, 180, 220
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7
  Pad 4 | Parameter: PROTEIN | Colour: Reddish | RGB: 220, 80, 180
    Value: Not provided (JSON value is null or absent) | Status: Color extracted; parameter label follows configured strip order; clinical value not assigned | Area: 7

Limitations: Results depend on lighting, camera, and calibration.
Parameter labels follow the configured top-to-bottom order; confirm this mapping for the exact strip model.
Clinical values remain unassigned until calibrated reference data is supplied.
Automated detection has not been validated on real strip photographs.