



Phantom QA results for / fbirn_out/ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc.xml, slice 34





[mean, SNR, SFNR] = [46137.7 137.5 133.6]






rdc = 9.2 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [1.622 1.671 1.752]
Smoothness(FWHM) in mm
0
1
2
3
4
0
10
20
30
40
50
60
70
80
90
100
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.116 2.163 2.222]
Smoothness(FWHM) in mm
0
1
2
3
4
0
10
20
30
40
50
60
70
80
90
100
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [0.721 1.007 1.153]
Smoothness(FWHM) in mm
Timepoint


Center of Mass in mm - X: [maxdisplacement drift] = [0.012 0.003]
cmassx
3dvolreg
Center of Mass in mm
Timepoint


Center of Mass in mm - Y: [maxdisplacement drift] = [0.053 -0.003]
cmassy
3dvolreg
Center of Mass in mm
Timepoint






Odd-even difference image

-13314	10829




image min: -13314, image max: 10829



Mean image




image min: 0, image max: 51310.7



Standard Deviation image




image min: 0, image max: 492.11



SFNR image

0 261.996

image min: 0, image max: 270.969



Acquisition parameters

dimensions	106x106x69x100
spacing	2mmx2mmx2mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	d10c7553-a968-4e
studyid	Stability_20261012
seriesnumber	35
runnumber	1
magneticfield	3

description	ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc
scandate	2026-10-12
scantime	12:29:01
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	2
frequencydirection	1
receivecoilname	HeadNeck_20_TCS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261012122859.222500
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexitimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	2
echotrainlength	80
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	RESEARCH
acquisitionduration	217.89500000000001

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	100
imagecomments	Not for diagnostic use, Unaliased MB3/ PE3/LB
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	69
rows	106
columns	106
bitsallocated	16
bitsstored	16
highbit	15
lossyimagecompression	00