




Phantom QA results for / fbirn_out/ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc.xml, slice 34









rdc = 6.1 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [1.592 1.647 1.775]
Smoothness(FWHM) in mm
4
3
2
1
0
0
10
20
30
40
50
60
70
80
90
100
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.092 2.139 2.238]
Smoothness(FWHM) in mm
4
3
2
1
0
0
10
20
30
40
50
60
70
80
90
100
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [0.883 1.075 1.224]
Smoothness(FWHM) in mm
Timepoint


Center of Mass in mm - X: [maxdisplacement drift] = [0.020 -0.011]
cmassx
3dvolreg
Center of Mass in mm
Timepoint


Center of Mass in mm - Y: [maxdisplacement drift] = [0.060 0.005]
cmassy
3dvolreg
Center of Mass in mm
Timepoint






Odd-even difference image




image min: -10941, image max: 9377



Mean image

0 53756.4


image min: 0, image max: 53788.3



Standard Deviation image

0 458.796


image min: 0, image max: 458.796



SFNR image

0 270.729

image min: 0, image max: 291.501



Acquisition parameters

dimensions	106x106x69x100
spacing	2mmx2mmx2mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	dc08a757-1251-4f
studyid	Stability_20261109
seriesnumber	35
runnumber	1
magneticfield	3

description	ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc
scandate	2026-11-09
scantime	12:34:16
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	2
frequencydirection	1
receivecoilname	HeadNeck_20_TCS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-20ch_func-bold_acq-cmrrstd_task-tsnrxacpc
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261109123414.440000
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexitimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	2
echotrainlength	80
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	RESEARCH
acquisitionduration	217.892

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	100
imagecomments	Not for diagnostic use, Unaliased MB3/ PE3/LB
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	69
rows	106
columns	106
bitsallocated	16
bitsstored	16
highbit	15
lossyimagecompression	00