




Phantom QA results for / fbirn_out/ses-20ch_func-bold_acq-siemensstd_task-tsnr.xml, slice 17









rdc = 10.0 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [2.171 2.298 2.648]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.467 2.579 2.871]
Smoothness(FWHM) in mm
0.5
1
1.5
2
2.5
3
3.5
4
4.5
5
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [1.175 1.666 1.954]
Smoothness(FWHM) in mm
Timepoint
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200


Center of Mass in mm - X: [maxdisplacement drift] = [0.117 -0.117]
cmassx
3dvolreg
Center of Mass in mm
Timepoint
106
106.5
107
107.5
0
20
40
60
80
100
120
140
160
180
200


Center of Mass in mm - Y: [maxdisplacement drift] = [0.126 0.118]
cmassy
3dvolreg
Center of Mass in mm
Timepoint
104
104.5
105
105.5
0
20
40
60
80
100
120
140
160
180
200






Odd-even difference image



image min: -342, image max: 388



Mean image

0 1460.77


image min: 0, image max: 1474.56



Standard Deviation image

0 9.7373


image min: 0, image max: 10.2884



SFNR image

0 302.63

image min: 0, image max: 308.557



Acquisition parameters

dimensions	68x68x36x200
spacing	3mmx3mmx3mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	dc08a757-1251-4f
studyid	Stability_20261109
seriesnumber	31
runnumber	1
magneticfield	3

description	ses-20ch_func-bold_acq-siemensstd_task-tsnr
scandate	2026-11-09
scantime	12:20:16
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	3
frequencydirection	1
receivecoilname	HeadNeck_20_TCS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-20ch_func-bold_acq-siemensstd_task-tsnr
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261109122014.135000
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexiimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	3
echotrainlength	33
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	PRODUCT
acquisitionduration	407.92500000000001

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	200
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	36
rows	68
columns	68
bitsallocated	16
bitsstored	12
highbit	11
lossyimagecompression	00