




Phantom QA results for / fbirn_out/ses-64ch_func-bold_acq-siemensstd_task-tsnr.xml, slice 17









rdc = 10.7 pixels
calc.
meas.
Relative std, %
10
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [2.072 2.141 2.210]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.406 2.458 2.521]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [1.283 1.388 1.480]
Smoothness(FWHM) in mm
3
2
1
0
-1
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Center of Mass in mm - X: [maxdisplacement drift] = [0.011 0.000]
cmassx
3dvolreg
Center of Mass in mm
103.5
103
102.5
102
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Center of Mass in mm - Y: [maxdisplacement drift] = [0.039 -0.021]
cmassy
3dvolreg
Center of Mass in mm
99.5
99
98.5
98
0
20
40
60
80
100
120
140
160
180
200
Timepoint






Odd-even difference image



image min: -501, image max: 363



Mean image

0 1338.75


image min: 0, image max: 1338.75



Standard Deviation image

0 12.7095


image min: 0, image max: 12.7095



SFNR image



image min: 0, image max: 530.64



Acquisition parameters

dimensions	68x68x36x200
spacing	3mmx3mmx3mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	847307d4-ec84-45
studyid	Stability_20261005
seriesnumber	4
runnumber	1
magneticfield	3

description	ses-64ch_func-bold_acq-siemensstd_task-tsnr
scandate	2026-10-05
scantime	11:00:30
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	3
frequencydirection	1
receivecoilname	HeadNeck_64_CS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-64ch_func-bold_acq-siemensstd_task-tsnr
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261005110028.707500
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexiimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	3
echotrainlength	33
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	PRODUCT
acquisitionduration	407.92500000000001

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	200
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	36
rows	68
columns	68
bitsallocated	16
bitsstored	12
highbit	11
lossyimagecompression	00