






Phantom QA results for / fbirn_out/ ses64ch_funcbold_acqcmrrstd_taskwar slice 34




old acqcmrrstd taskwarmup 2.xml [percent fluct (trend removed), drift, driftfit]





[mean, SNR, SFNR] = [31112.1 113.4 119.9]










Odd-even difference image

-7368 10914




image min: -7368, image max: 10914



Mean image




image min: 0, image max: 43233.3



Standard Deviation image




image min: 0, image max: 330.69



SFNR image



image min: 0, image max: 850.938



Acquisition parameters

dimensions	106x106x69x150
spacing	2mmx2mmx2mmx1[units?]
gap	0mmx0mmx0mmx0[units?]
scanner	AWP213018
psdname	
examnumber	0a0257d9-2388-4f
studyid	Stability_20250519
seriesnumber	2
runnumber	1-150
scanningsequence	GR/EP

sequencevariant	SK
magneticfield	3
description	ses-64ch_func-bold_acq-cmrrstd_task-warmup
scandate	2025-05-19
scantime	09:52:16
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	2
frequencydirection	1
acquisitionmatrix	106 106
bandwidth	192.284
receivecoilname	HeadNeck_64_CS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
ti	
protocolname	ses-64ch_func-bold_acq-cmrrstd_task-warmup
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20250519095214.065000
volumetricproperties	DISTORTED
scanoptions	PFP/FS/PER
mracquisitiontype	2D
sequencename	epfid2d1_106
angioflag	N
numaverages	1
imagingfrequency	123.253269
imagednucleus	1H
spacingbetweenslices	2
numphaseencodingsteps	106
echotrainlength	80
percentsampling	100
percentphasefieldofview	100

scannerserialnumber	213018
softwareversions	syngo MR XA61
triggertime	
beatrejectionflag	N
cardiacnumberofimages	1
inplane phase encoding direction	COL
variableflipangleflag	N
sar	0.01543863448386
contentqualification	RESEARCH
acquisitionnumber	1
instancenumber	1
numtemporalpositions	150
imagecomments	Not for diagnostic use, Unaliased MB3/PE3/LB
samplesperpixel	1
photometricinterpretation	MONOCHROME2
rows	106
columns	106
bitsallocated	16
bitsstored	16
highbit	15
lossyimagecompression	00
echonumber	0
pixelbandwidth	1814