



Phantom QA results for / fbirn_out/ses-32ch_func-bold_acq-siemensstd_task-tsnr.xml, slice 17





[mean, SNR, SFNR] = [896.9 115.7 125.3]






rdc = 9.1 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [2.140 2.219 2.296]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.421 2.498 2.569]
Smoothness(FWHM) in mm
0
1
2
3
4
5
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [1.141 1.450 1.679]
Smoothness(FWHM) in mm
Timepoint


Center of Mass in mm - X: [maxdisplacement drift] = [0.009 0.001]
cmassx
3dvolreg
Center of Mass in mm
Timepoint


Center of Mass in mm - Y: [maxdisplacement drift] = [0.026 -0.014]
cmassy
3dvolreg
Center of Mass in mm
Timepoint






Odd-even difference image



image min: -387, image max: 355



Mean image

0 1286.14


image min: 0, image max: 1286.14



Standard Deviation image

0 9.41512


image min: 0, image max: 9.97894



SFNR image

0 592.656

image min: 0, image max: 597.588



Acquisition parameters

dimensions	68x68x36x200
spacing	3mmx3mmx3mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	31ad03b6-d380-44
studyid	Stability_20261028
seriesnumber	19
runnumber	1
magneticfield	3

description	ses-32ch_func-bold_acq-siemensstd_task-tsnr
scandate	2026-10-28
scantime	11:06:47
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	3
frequencydirection	1
receivecoilname	Head_32
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-32ch_func-bold_acq-siemensstd_task-tsnr
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261028110645.690000
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexiimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	3
echotrainlength	33
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	PRODUCT
acquisitionduration	407.92500000000001

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	200
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	36
rows	68
columns	68
bitsallocated	16
bitsstored	12
highbit	11
lossyimagecompression	00