



Phantom QA results for / fbirn_out/ses-32ch_func-bold_acq-cmrrstd_task-tsnrxacpc.xml, slice 34





[mean, SNR, SFNR] = [29935.3 142.2 142.9]






rdc = 4.3 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [1.779 1.847 1.961]
Smoothness(FWHM) in mm
0
1
2
3
4
0
10
20
30
40
50
60
70
80
90
100
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [2.274 2.335 2.431]
Smoothness(FWHM) in mm
0
1
2
3
4
0
10
20
30
40
50
60
70
80
90
100
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [0.912 1.174 1.345]
Smoothness(FWHM) in mm
Timepoint


Center of Mass in mm - X: [maxdisplacement drift] = [0.015 -0.009]
cmassx
3dvolreg
Center of Mass in mm
Timepoint


Center of Mass in mm - Y: [maxdisplacement drift] = [0.051 -0.001]
cmassy
3dvolreg
Center of Mass in mm
Timepoint






Odd-even difference image




image min: -7039, image max: 6852



Mean image

0 40827.6


image min: 0, image max: 40827.6



Standard Deviation image

0 260.966


image min: 0, image max: 263.495



SFNR image

0 661.956

image min: 0, image max: 661.956



Acquisition parameters

dimensions	106x106x69x100
spacing	2mmx2mmx2mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	a0f8a673-da21-44
studyid	Stability_20261019
seriesnumber	28
runnumber	1
magneticfield	3

description	ses-32ch_func-bold_acq-cmrrstd_task-tsnrxacpc
scandate	2026-10-19
scantime	12:33:03
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	2
frequencydirection	1
receivecoilname	Head_32
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-32ch_func-bold_acq-cmrrstd_task-tsnrxacpc
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261019123301.005000
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexiimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	2
echotrainlength	80
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	RESEARCH
acquisitionduration	217.89500000000001

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	100
imagecomments	Not for diagnostic use, Unaliased MB3/ PE3/LB
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	69
rows	106
columns	106
bitsallocated	16
bitsstored	16
highbit	15
lossyimagecompression	00