




Phantom QA results for / fbirn_out/ses-64ch_func-bold_acq-siemensbasic_task-tsnr.xml, slice 14









rdc = 10.6 pixels
calc.
meas.
Relative std, %
1
0.1
0.01
1
10
ROI full width, pixels



Smoothness(FWHM) in mm - X: [min mean max] = [2.360 2.444 2.573]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint



Smoothness(FWHM) in mm - Y: [min mean max] = [1.904 2.124 2.248]
Smoothness(FWHM) in mm
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200
Timepoint


Smoothness(FWHM) in mm - Z: [min mean max] = [1.258 1.619 2.111]
Smoothness(FWHM) in mm
Timepoint
0
1
2
3
4
0
20
40
60
80
100
120
140
160
180
200


Center of Mass in mm - X: [maxdisplacement drift] = [0.007 0.002]
Center of Mass in mm
Timepoint
cmassx
3dvolreg
102.5
103
103.5
104
0
20
40
60
80
100
120
140
160
180
200


Center of Mass in mm - Y: [maxdisplacement drift] = [0.039 -0.031]
Center of Mass in mm
Timepoint
cmassy
3dvolreg
98.5
99
99.5
100
0
20
40
60
80
100
120
140
160
180
200






Odd-even difference image



image min: -237, image max: 243



Mean image

0 1327.25


image min: 0, image max: 1327.25



Standard Deviation image

0 7.63845


image min: 0, image max: 7.63845



SFNR image



image min: 0, image max: 950.514



Acquisition parameters

dimensions	68x68x30x200
spacing	3mmx3mmx3mmx2000ms
gap	0mmx0mmx0mmx0ms
scanner	AWP213018
psdname	
examnumber	60e5c74d-1732-4f
studyid	Stability_20261102
seriesnumber	8
runnumber	1
magneticfield	3

description	ses-64ch_func-bold_acq-siemensbasic_task-tsnr
scandate	2026-11-02
scantime	11:54:42
tr	2000
te	30
operator	
flipangle	79
prescribedslicespacing	3
frequencydirection	1
receivecoilname	HeadNeck_64_CS
transmitcoilname	Body
institution	THE UNIVERSITY OF PENNSYLVANIA
examdescription	ACTIVE KIRWAN
scanner	AWP213018
scannermanufacturer	Siemens Healthineers
scannermodelname	MAGNETOM Cima.X
protocolname	ses-64ch_func-bold_acq-siemensbasic_task-tsnr
institutionaddress	Grays Ferry Ave. BLDG 250 STE 140 3401,Philadelphia,69 - KEYSTONE,US,19146
institutionaldepartmentname	MindCORE Neuroimaging Suite 140
acquisitiondatetime	20261102125440.560000
frametype	ORIGINAL
pixelpresentation	MONOCHROME
volumetricproperties	DISTORTED
volumebasedcalculationtechnique	NONE
complexiimagecomponent	MAGNITUDE
acquisitioncontrast	UNKNOWN
mracquisitiontype	2D
numaverages	1
spacingbetweenslices	3
echotrainlength	68
percentsampling	100
percentphasefieldofview	100
scannerserialnumber	213018
softwareversions	syngo MR XA61
inplanephaseencodingdirection	COLUMN
variableflipangleflag	N
contentqualification	PRODUCT
acquisitionduration	405.91200000000003

resonantnucleus	1H
applicablesafetystandardagency	IEC
acquisitionnumber	1
instancenumber	1
numtemporalpositions	200
samplesperpixel	1
photometricinterpretation	MONOCHROME2
numberofframes	30
rows	68
columns	68
bitsallocated	16
bitsstored	12
highbit	11
lossyimagecompression	00