Dual-ligand screen / Setup

48 pairs, then 22 pairs + L17 reference. One 48 x 2 mL Para-Dox, reused with **71 fresh reaction vials**. Manual screw caps; no stir bars. Keep reaction and LC maps separate.

Before the clock

- Fit **p300 single GEN2 left; p300 multi GEN2 right**. Single only for 13 mm-pitch reaction vials. Check block/latch, offsets, cap height <=45 mm and tips >1 mm above bottoms. Simulate released Python locally.
- Use matching `dosing.csv`, `stock_preparation.csv`, `deck_loads.csv`, `lc_sequence.csv` and `round2_predose.csv`. A software colleague handles release and LC imports. Stocks and initial scoring are outside the clock.
- Have 71 reaction vials/caps, **two final 96-well LC plates** matching the VWR 2.2 mL geometry, compatible pierceable seals and fresh tips; standards separately. Confirm plate fit on shaker AND autosampler, needle depth and **full 1.7-min LC cycle**.

Prepare stocks

Stock	Bulk preparation / use
SM + mesitylene	**2.0 mL: 1 M SM +100 mM IS in toluene.** 322.32 mg SM +24.04 mg IS at 100% assay; make to FINAL volume, 4 mL capped vial. Dose **25 µL**; retain solvent.
Ligands / Pd	**Ligands 20 mM; Pd 40 mM in toluene.** CSV masses/reserves; prepare ligands in glovebox. Working tubes **2 mL**, max 1.95 mL.
RockPhos L03	**800 µL, 20 mM:** 7.50 mg pure equivalent; correct for assay. **10 mg available; maximum 4 R2 wells.**
Quench	**100 mL, 9:1 ACN/water +0.15% formic acid:** 150 µL acid, make to volume. No IS.
LC diluent	**125 mL, 1:1 water/ACN +0.1% formic acid:** 125 µL acid, make to volume. No IS.

Correct masses for assay. **Load stage-specific CSV volumes, not bulk preparations.** Refill R2 from reserves; ordinary plastic source racks are sufficient.

Deck

Slot	Contents
1 / 4	2 mL working tubes: ligands at CSV addresses; **SM/IS in slot 4 D2, 1.31 mL for R1**. L07 absent
2	A2 Pd; quench A4/A6/A9; diluent A5/A7/A10/A11/A12. A1/A3/A8 unused
3 / 10	Reaction block: staff move it by hand at App prompts; pipetting / heater-shaker
6	Final 96-well 2.2 mL LC plate: fill here; move to 10, then directly to HPLC
5 / 8 / 9	Single tips / multi tips / multi tips. Keep 7 and 11 empty

Calibration during hold one - 14 injections

Each SM and product: blank + **10, 50, 150, 350, 800 µM**, plus independent **666.7 µM** check. **Mesitylene stays at 66.7 µM in every standard.** Use matched matrix and calibration recipes. HTEL exports UV areas; Python calibrates and calculates yields. The software colleague handles method/RT settings and pilot checks.

Dual-ligand screen / Run

Operator / start: _____ LC method / channel: _____

Follow App pauses. Hand-tighten reaction-vial caps before heating/shaking; the LC plate shakes uncovered. Record actual times and deviations.

Step	Action
1 / SM + IS	Dose **25 µL combined stock** into 48 open vials. Keep the toluene. No separate IS addition.
2 / Reaction	Add **75 µL ligand A +75 µL ligand B +75 µL Pd =250 µL total**. Hand-cap; move/latch as prompted. **95 °C, 500 rpm, 4 h.**
3 / During hold	Run calibration. Hand-dose **23 labelled NEW vials with 25 µL combined SM/IS stock each** from off-deck reserve. **Cap promptly; retain toluene.** No catalyst yet.
4 / Cool	Move the sealed block off the heater. Cool vials before opening; App also waits for module <=30 °C. Load workup sources.
5 / Quench	Add **500 µL quench** per vial. Hand-recap ALL. Shake **1500 rpm, 2 min**, heater off; return to slot 3.
6 / Sample	Let bubbles settle after shaking; uncap the indicated reaction vials. Robot adds **20 µL extract to 980 µL diluent** with fresh sample tips.
7 / LC handoff	Recap reaction vials. Move the completed LC plate **6** → **cooled shaker 10** at App prompt. Robot shakes **UNCOVERED, heater OFF, 1000 rpm, 2 min** (locally confirmed splash-free speed). Move **10** → **HPLC**, apply compatible seal. Inject directly from this plate; **no manual dilution/mixing or vial transfers**.
8 / R2 setup/select	During R1 LC, archive sampled R1 vials; fit 23 capped predosed vials into the SAME block. After all 48 results pass QC, select 22 pairs +L17 reference; **RockPhos at most 4 wells**.
9 / R2	Refill R2 deck loads; **uncap all 23 NEW vials. No SM/IS redose.** Add catalysts; reference uses **150 µL L17**, no B. Cap; repeat 4 h reaction/workup. One new LC plate; **H3 is diluent only: do not inject.**

Stop the clock after **23 final injections and quantitative analysis**. Give the pipeline operator robot JSON logs, raw LC/MS files, stock records, actual times and deviations for the private return ZIP.

If a transfer looks wrong

Pause for layers, solids at aspiration, persistent bubbles, air uptake or tip contact. Repeated aspiration does not fix solubility. Reuse tips only for approved same-stock/noncontact dosing; samples always get fresh tips. Replace an uncertain dilution with a fresh documented one; do not guess a top-up.

12-hour operating target: 8 h reaction holds +120.7 min LC; **119.3 min for all other tasks and contingency**. Calibration/R2 SM predosing overlap hold one; R2 block changeover can overlap R1 LC after sampling. Confirm actual cycle/handling times before release. **Four LC-plate moves; no R1 split.** Full procedure and local commands.