

Major

- Right now the bin sizes for templates are an hour - they need to be variable based on the signal (Deep)
- Need a graph (you can remove it) on the ECG noise marker with small boxes 0.1s apart and larger boxes 0.5s apart (Mohamed)
- Unsure what to do with the PPG is shifted with respect the ECG because then my shifting method for making the ECG not overlap with the PPG in the template marking doesn't work (Me)

Minor

- Add a std light gray outline around the respiratory mark (DK)
- Add a simple icon to the app so users know what is being run (DK)