MSK Annotation Suite

Clinical annotation toolkit for local MSK — FFOL, Insall-Scuderi, Patellar TA, and Sulcus Angle with DICOM support and structured CSV export.

Table of Contents

- [MSK Annotation Suite](#)
- [Table of Contents](#)
- [Running the Code](#)
- [Deploying to Render](#)
- [What We Export & Why](#)
- [The Golden Thread: How Coordinates Stay Reproducible](#)
- [CSV Export Architecture](#)
  - [1. Session Annotation CSV (`sessions/{id}/annotations.csv`)](#)
  - [Core Columns (12 columns)](#)
  - [Measurement Columns (variable columns)](#)
  - [`measurementType` Values and Their Geometry](#)
  - [2. Protocol Measurement CSV (`sessions/{id}/protocolMeasurements.csv`)](#)
  - [Core Protocol Measurement Columns](#)
  - [Cross-Protocol Annotation Summary](#)
  - [Patient Table](#)
  - [Protocol A: Coronal Alignment](#)
  - [Example: A Complete FFOL Protocol Row](#)
- [Measurement Protocols & Required Annotations](#)
  - [Protocol Completion Tracking](#)
  - [Post-Fix Items for MI](#)
  - [DICOM Spatial Metadata](#)
- [DICOM Data Architecture](#)
  - [Volume Structure](#)
  - [Coordinate Systems](#)
- [Machine Learning: Predicting Measurements from DICOM](#)
  - [Problem Formulation](#)
  - [Given: Three co-registered 3D DICOM volumes — sagittal V_{sag}, coronal V_{cor}, axial V_{ax} — each of shape (D, H, W) single-channel, with known pixel spacing $(\Delta x, \Delta y, \Delta z)$ mm and DICOM origin.](#)
  - [Predict: One or more clinical measurements. There are two output granularities:](#)
  - [Landmark-level](#)
  - [Value-level](#)
  - [Recommendation: Landmark-level prediction](#)
  - [Input — Output Mapping](#)
  - [Protocol 1: TT-TG (Axial)](#)
  - [Protocol 2: Insall-Scuderi (Sagittal)](#)
- [Reference Line System](#)
  - [Machine Learning: Predicting Measurements from DICOM](#)
  - [Problem Formulation](#)
  - [Input — Output Mapping](#)
  - [Protocol 1: TT-TG (Axial)](#)
  - [Protocol 2: Insall-Scuderi (Sagittal)](#)
- [Measurement Protocols & Required Annotations](#)
  - [Protocol Completion Tracking](#)
  - [Post-Fix Items for MI](#)
  - [DICOM Spatial Metadata](#)
- [Coordinate Systems](#)
- [Reference Line System](#)
- [Machine Learning: Predicting Measurements from DICOM](#)
- [Problem Formulation](#)
- [Input — Output Mapping](#)
- [Protocol 1: TT-TG (Axial)](#)
- [Protocol 2: Insall-Scuderi (Sagittal)](#)

Running the Code

```
npm install
npm run dev
```

Note: The project uses Vite. If you don't have it, install it: `npm install -g vite`

Deploying to Render

This project is configured for one-click deployment on Render as a static site. The `render.yaml` will be detected automatically.

1. Push the repo to GitHub/GitLab.
2. On Render, create a new **Static Site**.
3. Connect your repository.
4. Render will auto-detect the `render.yaml` config, or set manually:
   - **Build Command**: `npm install && npm build`
   - **Publish Directory**: `dist`

What We Export & Why

The tool produces **two complementary CSV formats** that together form a complete ground-truth dataset for ML training. They were inflated purposes:

CSV	Purpose	Granularity	Best for
Session Annotation CSV	One row per annotation (line, point, angle)	One row per annotation	Landmark detection training
Protocol Measurement CSV	One row per measurement, grouped by patient/protocol	One row per measurement	Value regression training + clinical audit

The Golden Thread: How Coordinates Stay Reproducible

The most critical design decision: **every coordinate is stored in both voxel and world space**. Voxel indices are meaningless without the corresponding volume's `imageSize`, `voxelSpacing`, and `imageOrientationPatient`.

```
imageIndex = (
  x: sagittal_slice_index,
  y: coronal_slice_index,
  z: axial_slice_index
)
worldPosition = (
  x: -142.3,   // LPS: patient's left
  y: -87.6,    // LPS: anterior
  z: 1203.4    // LPS: superior
)
```

This is captured **at the moment the annotation is drawn** — before the user ever moves their window. Combined with DICOM spatial metadata from the volume header, coordinates are always reconstructable.

CSV ex — Patient mm in practice (for a 256×256 image displayed at 400×400 px with 0.8 mm/pixel spacing):

```
measurementType = 'FFOL_TA'
subject_id = 'ANON4471'
study_id = '20240315'
image_id = '1.2.840.113619.2.55.3.604688119.969.1234567890.123'
imageIndex = (x: 128, y: 128, z: 87)
worldPosition = (x: -142.3, y: -87.6, z: 1203.4)
imageOrientationPatient = (1, 0, 0, 0, 1, 0)
pixelSpacing = (0.8, 0.8)
sliceThickness = 3.0
value = 12.4
unit = 'mm'
side = 'left'
visibleDeg = 41
angleDeg = 38.2
confidence = 'high'
```

Critically, the affine submission contains **no** dimension (`x`, `y`, `z`) or offset columns, but is essential for point and perpendicular measurements. It is stored separately in the protocol CSV.

CSV Export Architecture

1. Session Annotation CSV (`sessions/{id}/annotations.csv`)

Every click, every line, every measurement — raw and unprocessed. This is the ML training dataset. Each row is a single drawing action by the clinician. One is the most granular training layer for landmark prediction models.

Core Columns (12 columns)

Column	Type	Example	Meaning
`annotationId`	string	`a3f2c1d4-...`	Globally unique per drawing
`patientId`	string	`ANON4471`	DICOM PatientID tag (`0010,0020`)
`studyId`	string	`20240315`	Which exam (`0020,0010`)
`imageInstance`	string	`1.2.752.24...`	DICOM Series Description — identifies the specific image
`imageIndex`	integer	`87`	Anatomical slice of the volume
`measurementType`	enum	`perpendicular`	Drawing tool used. (`line`, `distance`, `point`, `angle`, `perpendicular`)
`value`	string	`12.4` or `""`	In viewport measurement label (CSS pixel space) or empty for points
`unit`	string	`mm`	Unit suffix parsed from the label
`visibleDeg`	int	`41`	If based slice within the volume where annotation was
`annotatorId`	string	`DC`, `SarahS`	Clinician name
`annotationDateTime`	string	`2024-03-15T14:22:08Z`	Clinician name for audit trail
`timeSpent`	ISO 8601	`PT1M30S`	Time spent on this annotation

Extended ML Columns (20 columns)

Column	Type	Meaning
`label`	string	Human-readable label, e.g. `"Femoral Canal Landmark"`
`worldPosition`	string	Which protocol step this fulfills (`canalTip`, `tibialSpineMidpoint`, etc.)
`groupId`	string	Links annotations of the same protocol run (e.g. `"FFOL_TA"`)
`pixelSpacing`	string	Unique annotations number in a group
`points_count`	int	Number of points in this annotation's geometry
`dX_, dY_, dZ_`	float	Up to 4 Hounsfield Unit values at sampled
`WorldPoint_X, WorldPoint_Y, WorldPoint_Z`	float	Reserved/reference line endpoints (in volume), reserved for future use
`WorldPoint_points_json`	JSON	Full `[{x,y,z}, ...]` array for programmatic consumption
`WorldPoint_points_csv`	JSON	Full geometry point array

Why two representations? The flattened `dX_, dY_, dZ_` columns let you filter/sort in Excel. The `points_json` column preserves the full geometry for programmatic reconstruction. Both are written on every export.

`measurementType` Values and Their Geometry

Type	Points Count	Typical Use	Example
`point`	1	Joint line placement, landmark	Femoralline point
`line`	2	Complete line, patella axis	Posterior femoral condyle line
`distance`	2	Length measurements	Patella length (p10, tendon length, etc.)
`angle`	3	Angle between two lines (vertex at point 2)	Any 3-point angle measurement
`perpendicular`	2	Bisect off a reference line	Perpendicular to a reference line

What makes this ML-ready: Every row is a self-contained training example. The `dX_, dY_, dZ_` columns are the regression targets for a landmark detector. The `label` and `worldPosition` columns are the classification targets. The `imageInstance` tells you which slice to load. And the `pixelSpacing` tells you how to convert pixel distances to mm.

2. Protocol Measurement CSV (`sessions/{id}/protocolMeasurements.csv`)

Clinical audit, reporting, and ML validation. This CSV provides ground-truth values for clinical measurements, with full provenance tracking. It's designed for regression models and exports the final clinical values alongside all intermediate data — **physical mm**, not pixel units.

Core Columns for Machine & Protocol Runs

The exporter uses a **3-pass inference system** to determine which annotations belong to which protocol steps:

- **Pass 1 — Explicit `groupId`:** Annotations with a `groupId` field are matched directly to protocol runs. This is the most reliable method.
- **Pass 2 — Label prefix matching:** Annotations whose `label` starts with a protocol prefix (e.g. `"FFOL_TA"`) are matched by label prefix. This is the fallback for annotations created before `groupId` was added.
- **Pass 3 — Label + patientId matching:** Annotations whose `label` contains a protocol keyword (e.g. `"Femoral Canal Landmark"`) are matched by label + patientId. This is the most permissive pass.

Unmatched rows are still exported — they just have empty `protocolId` and `protocolName` columns. This preserves annotations that don't fit any protocol, such as free-form measurements or annotations from protocols that haven't been implemented yet.

Cross-Protocol Annotation Summary

A critical feature: **annotations shown for one protocol can satisfy steps in another.** For example, the posterior femoral condyle line drawn during TT-TG is automatically used by Insall-Scuderi. The export tracks this:

- Groups measurements by `patientId` + `groupId` (or `protocolId`)
- Counts how many annotations each protocol step received
- Reports how many annotations were shared (used by) more than one protocol
- Shadow columns (`sessionId`) provide the the borrowing group (the original study) in its own group

This makes a complete row and it's present in the CSV and also the complete row **once**.

Record Types

Each row has a `recordType` field that indicates its role:

recordType	Meaning	Key Columns Populated
`protocol_run`	Final clinical measurement for one protocol	`value`, `unit`, `protocolId`, `protocolName`, `completedAt`
`step_measurement`	Individual annotation within a protocol run	`measurementId`, `stepId`, `dX_, dY_, dZ_` (in mm), `pixelSpacing`
`raw_measurement`	Annotation that doesn't belong to any recognized protocol	All geometry columns, no protocol metadata

Full Column Reference

Column	Source	Description
`recordType`	generated	`protocol_run`, `step_measurement`, or `raw_measurement`
`patientId`	DICOM	Patient identifier
`sessionId`	session	Clinician name at export time
`exportedAt`	session	Clinician name at export time
`frameOfRef`	annotation	`left` or `right`
`imageInstance`	annotation	Frame sequence description
`groupId`	annotation	Anatomical plane
`protocolId`	annotation	`"FFOL_TA"`, `"patellar"`, `"patellar_TT"`, `"sulcus_angle"`, `"sulcus_angle_tka"`, `"canal_deviation"`
`protocolName`	annotation	Human-readable name
`groupId`	annotation	Groups annotations of the same protocol run
`measurementId`	annotation	UUID of the individual annotation
`stepId`	annotation	Protocol step ID (`"patellaTip"`, `"femoralCondyleLength"`, etc.)
`stepLabel`	annotation	Human-readable step name
`groupId`	annotation	Groups annotations of the same protocol run
`measurementType`	annotation	Actual drawing type (`line`, `distance`, `point`, `angle`, `perpendicular`)
`visibleDeg`	annotation	If based slice number
`pointsTotal`	computed	Number of points in this annotation
`rawValue`	computed	Final clinical value (in mm or degrees)
`rawUnit`	computed	`mm`, `°`, or empty for ratios
`rawUncertainty`	computed	Human-readable, e.g. `"TT-TG = 12.4 mm"`
`interpretation`	computed	Clinical interpretation, e.g. `"Elevated (>20 mm)"` or `"Lateralized (<-10 mm)"`
`measurementCount`	computed	Number of annotations used in this protocol run
`measurementId`	computed	Human-readable label
`measurementStep`	computed	Per-annotation computed value (distance in mm or angle in °)
`measurementUnit`	computed	Geometry endpoints in physical mm
`imageOrientation`	computed	Up to 4 points (2 annotations = 12 columns. **Coordinates are in physical mm**, not pixels. Empty if fewer than 2 annotations.
`dX_, dY_, dZ_`	computed	`pointOneStartX_mm`, `pointOneStartY_mm`, `pointOneStartZ_mm`, `pointOneEndX_mm`, `pointOneEndY_mm`, `pointOneEndZ_mm`, `pointTwoStartX_mm`, `pointTwoStartY_mm`, `pointTwoStartZ_mm`, `pointTwoEndX_mm`, `pointTwoEndY_mm`, `pointTwoEndZ_mm`
`imageOrientationPatient`	computed	Absolute value of $(\Delta x, \Delta y, \Delta z)$ — the anatomical displacement vector in physical mm
`sagittalDistance`	computed	$\sqrt{\Delta x^2 + \Delta y^2 + \Delta z^2}$ — the straight-line distance between the two endpoints
`SagittalDistance_mm`	computed	Same as `imageOrientationPatient` (kept for compatibility)
`SagittalDistance_mm`	duplicate	Same as `SagittalDistanceMm` (legacy compatibility)
`SagittalDistanceMm`	duplicate	Same as `SagittalDistanceMm` (legacy compatibility)
`pixelSpacing`	computed	Δx — the anatomical displacement in the sagittal plane (left–right)
`pixelSpacing`	duplicate	Δy — the anatomical displacement in the coronal plane (anterior–posterior)
`pixelSpacing`	duplicate	Δz — the anatomical displacement in the axial plane (superior–inferior)

Example: A Complete TT-TG Protocol Row

Here's a patient `ANON4471` and `20240315` exam, TT-TG was run by a TT-TG measurement. The complete row was drawn, then two perpendicular branches for groove and reference:

```
recordType,patientId,sessionId,exportedAt,frameOfRef,groupId,measurementId,stepId,stepLabel,protocolId,protocolName,rawValue,rawUnit,rawUncertainty,interpretation,measurementCount,measurementStep,measurementUnit,imageOrientation,dX_,dY_,dZ_,pointOneStartX_mm,pointOneStartY_mm,pointOneStartZ_mm,pointOneEndX_mm,pointOneEndY_mm,pointOneEndZ_mm,pointTwoStartX_mm,pointTwoStartY_mm,pointTwoStartZ_mm,pointTwoEndX_mm,pointTwoEndY_mm,pointTwoEndZ_mm,imageOrientationPatient,sagittalDistance,SagittalDistance_mm,SagittalDistanceMm,pixelSpacing,pixelSpacing,pixelSpacing
protocol_run,ANON4471,DC,2024-03-15T14:30:00Z,left,,,,,TT_TG,TT-TG,12.4,mm,"TT-TG = 12.4 mm",Elevated (>20 mm),2,,,,,
```

```
# Protocol 4: Saliva Angle (Axial)

Input: Axial scanner
Output: 3D ear contact angle vector (θ, φ, r)
1. Segment Trochlear groove boundary
2. Identify deepest point
Saliva Angle = angle(mandibular notch, fossa, superior of C.M)

# Protocol 5: Saliva Angle 3 cm (Cross-Plane)

Input: Sagittal volume (x,y,z) Axial volume (x,z)
Step 1: Segment
Step 2: Identify Trochlear point (the point at which y=y)
Step 3: Create plane at z=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 4: Create plane at x=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 5: Create plane at y=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 6: Create plane at x=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 7: Create plane at y=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 8: Create plane at z=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 9: Create plane at x=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 10: Create plane at y=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 11: Create plane at z=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 12: Create plane at x=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 13: Create plane at y=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step 14: Create plane at z=(Centroid(Mandible) + Cos(Saliva Angle)*3 cm)
Step
```

```
    'tt_tg':              0.5,   # medium — perpendicular branch
    'insall_salvati':     0.5,   # medium — tendon soft tissue
    'caton_deschamps':    0.3,   # hard — articular surface boundary
    'sulcus_angle_3cm':   0.3,   # hardest — cross-plane inference
}

# Phase schedule: linearly ramp up weights for harder protocols
for epoch in range(num_epochs):
    ramp = min(1.0, epoch / 60) if epoch < 60 else 1.0
    active_weights = {
        k: v * ramp if v < 1.0 else v for k, v in protocol_weights.items()
    }
```

Full Training Loop

```
# Phase 1: DINOv2 frozen, train heatmap head only
#   ~30 epochs, LR=1e-3, AdamW, cosine schedule

# Phase 2: Unfreeze encoder, full fine-tuning with curriculum
#   ~70 epochs, LR=1e-4, AdamW, cosine schedule with warmup

# Phase 3: Add clinical value loss, fine-tune end-to-end
#   ~50 epochs, LR=5e-5

optimizer = torch.optim.AdamW(model.parameters(), lr=1e-4, weight_decay=1e-4)
scheduler = torch.optim.lr_scheduler.CosineAnnealingWarmRestarts(
    optimizer, T_0=20, T_mult=2
)

# Use mixed precision for memory efficiency
scaler = torch.cuda.amp.GradScaler()

for epoch in range(num_epochs):
    for batch in dataloader:
        with torch.cuda.amp.autocast():
            pred = model(batch['v_sag'], batch['v_cor'], batch['v_ax'])
            loss, loss_log = criterion(pred, batch['targets'])
        scaler.scale(loss).backward()
        scaler.step(optimizer)
        scaler.update()
        scheduler.step()
```

Evaluation Metrics

Metric	Formula	Target
Landmark Error (mm)	$\frac{1}{K}\sum_k \lvert\hat{p}_k - p_k\rvert \cdot s$ where s is pixel spacing	< 2 mm (inter-rater variability)
TT-TG MAE	$\lvert\widehat{\text{TT-TG}} - \text{TT-TG}\rvert$	< 2 mm
Angle MAE	$\lvert\hat{\theta} - \theta\rvert$ for Patellar Tilt, Sulcus Angle	< 3°
Ratio MAE	$\lvert\hat{r} - r\rvert$ for Insall–Salvati, Caton–Deschamps	< 0.10
Slice Accuracy	% of cases where predicted slice = ground-truth slice	> 90%
Clinical Agreement	% of predictions in same clinical category (normal/borderline/abnormal)	> 85%
Hausdorff Distance (mm)	$\max\left(\sup_{p\in P}\inf_{g\in G}\lvert p-g\rvert, \sup_{g\in G}\inf_{p\in P}\lvert g-p\rvert\right)$ between predicted and ground-truth landmark sets	< 5 mm
Expected Calibration Error (ECE)	$\sum_{m=1}^{M}\frac{\lvert B_m\rvert}{n}\lvert\text{acc}(B_m) - \text{conf}(B_m)\rvert$	< 0.05

Why these additions matter:

- **Hausdorff Distance** catches cases where a secondary heatmap mode is placed far from the true landmark — MSE on heatmaps doesn't penalize this heavily, but it produces clinically wrong measurements.
- **ECE (Expected Calibration Error)** measures whether the model's stated confidence matches its empirical accuracy. A well-calibrated model that says "I'm 90% confident this landmark is here" is actually correct 90% of the time. This is **required for FDA 510(k) submissions** and is conspicuously missing from most medical AI evaluation.

```
from scipy.spatial.distance import directed_hausdorff
def hausdorff(pred_landmarks, gt_landmarks):
    return max(directed_hausdorff(pred_landmarks, gt_landmarks)[0],
               directed_hausdorff(gt_landmarks, pred_landmarks)[0])

def expected_calibration_error(confs, accs, n_bins=10):
    bins = np.linspace(0, 1, n_bins + 1)
    ece = 0.0
    for i in range(n_bins):
        mask = (confs >= bins[i]) & (confs < bins[i+1])
        if mask.sum() > 0:
            ece += (mask.sum() / len(confs)) * abs(accs[mask].mean() - confs[mask].mean())
    return ece
```

Inter-rater baseline: Knee MRI measurements have reported inter-rater ICC of 0.80–0.95 depending on protocol. The model should target performance within the clinician inter-rater range.

Dataset Requirements

For training a clinically useful model:

Phase	Minimum	Recommended
Feasibility (ViT regression)	100 annotated studies	300
Landmark heatmap model	300 annotated studies	500+
Clinical validation	50 held-out studies	100 held-out
External validation	30 studies from different scanner	50+

Each "study" = one patient knee (3 DICOM series) with all 6 protocol measurements annotated.

Clinical Gaps & Production Considerations

Annotation Versioning

The CSV exports currently have `annotatorEmail` and `timestamp`, but no `annotationVersion` or `correctionOf` field. Clinicians will revise annotations — you need a way to track which version was used to train each model. **Add these columns to both CSV exports:**

Column	Type	Description
`annotationVersion`	integer	Monotonically increasing; 1 = initial, 2+ = revision
`correctionOf`	UUID or null	If this is a correction, points to the previous `annotationId`
`modelVersion`	string or null	Which model version (if any) generated a suggested annotation that the clinician then corrected

This enables: (a) excluding superseded annotations from training, (b) measuring annotation drift over time, (c) active-learning loops where model predictions are clinician-corrected and fed back.

Two-Stage Inference for Sulcus Angle 3 cm

The Sulcus Angle 3 cm protocol is **cross-plane** and requires two-stage inference — the current architecture sketch doesn't accommodate this properly:

```
Stage 1 (Sagittal):
  Input:  V_S (full sagittal volume)
  Output: Joint line point J = (x_j, y_j) on sagittal slice s_sag
  Compute: Axial slice index s* using DICOM affine:
    patient_3d = sag_affine.voxelToPatient((y_j, x_j, s_sag))
    s* = axial_affine.patientToVoxel(patient_3d).slice + 30mm_offset

Stage 2 (Axial, at slice s*):
  Input:  V_A[:, s*, :, :] (single axial slice at s*)
  Output: Medial condyle M, sulcus S, lateral condyle L
  Compute: Sulcus angle from M, S, L
```

Implementation note: The DICOM affine transforms (`dicomAffine.ts`) already implement `voxelToPatient` and `patientToVoxel`. Port these to PyTorch as differentiable operations so the two-stage pipeline can be trained end-to-end (the slice selection is a hard argmax — use straight-through Gumbel-Softmax or REINFORCE for gradient estimation).

MONAI Integration

Replace custom preprocessing with MONAI's production pipeline:

```
from monai.transforms import (
    Compose, LoadImageD, ScaleIntensityD, SpatialPadD,
    RandRotateD, RandZoomD, RandGaussianNoiseD,
)
from monai.data import Dataset, DataLoader

transforms = Compose([
    LoadImageD(keys=['v_sag', 'v_cor', 'v_ax']),
    ScaleIntensityD(keys=['v_sag', 'v_cor', 'v_ax']),  # Z-score per-volume
    SpatialPadD(keys=['v_sag', 'v_cor', 'v_ax'], spatial_size=(256, 256)),
    RandRotateD(keys=['v_sag', 'v_cor', 'v_ax'], range_x=0.17, prob=0.5),  # ±10°
    RandZoomD(keys=['v_sag', 'v_cor', 'v_ax'], min_zoom=0.9, max_zoom=1.1, prob=0.5),
])
```

Next Steps

1. **Add `annotationVersion` / `correctionOf` / `modelVersion` columns** to both CSV exports in the UI
2. **Export training data** using the existing CSV export buttons
3. **Validate landmark consistency** — have 2+ clinicians annotate the same 20 studies, compute inter-rater ICC and Hausdorff distance between annotators (this becomes your target performance ceiling)
4. **Set up MONAI + PyTorch training environment** (replaces custom preprocessing)
5. **Train the DINOv2 ViT baseline** (Option B) first — quick to implement, establishes a performance floor
6. **Train the HRNet 2.5D landmark model** (Recommended) with Wing Loss, cross-plane attention, and curriculum learning
7. **Implement two-stage inference** for Sulcus Angle 3 cm (sagittal → affine → axial pipeline)
8. **Calibration study** — measure ECE, apply temperature scaling if needed
9. **Clinical validation study** comparing model vs. clinician measurements on held-out data

---

License

PolyForm Noncommercial License 1.0.0 — free for personal, educational, academic research, and noncommercial clinical use. Commercial use requires a separate license. See LICENSE for full terms.