

**TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS**

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010134	GROUP #	PHC297
-----------	---------------	---------	--------

ITEM DESCRIPTION	CHAIR LAB	اسم البند	كرسي مختبر
------------------	-----------	-----------	------------

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	TYPE	Gas Piston / Spring	
2	HEIGHT 42 to 55 Cm. ADJUSTABLE	Yes	
3	FRAME	Stainless Steel/ Aluminum, specify	
4	SWIVEL	Yes	
5	MUST HAVE ADEQUATE BACK SUPPORT	Yes	
6	SEAT SURFACE SOFT	Yes	
7	BASE FRAM	Anti-Tilt	
8	FOOTREST: RING TYPE	Yes	
9	CASTOR HEAVY DUTY	Yes (Five), non-conductive	
10	OTHER SPECIFICATIONS		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها