Kingdom of Saudi Arabia Ministry Of Health General Directorate of Equipment Supplies Department Of Specs. & Development



المملكة العربية السعودية وزارة الصحة إدارة التجهيزات الطبية إدارة المواصفات والتطوير

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية									
مراكز الرعاية الصحية									
445500040000								ODOUD # DUOLOO	
Item Code 44552600100		4455260010098						GROUP # PHC186	
ITEM DE	SCRIPTION	TABLE EXAM - GYN	ECOLOG'	Υ			فحص نسائية	اسم البند طاولة	
MODEL OFFERED			MANUFACTURER			COUNTRY OF ORIGIN			
#	TECHNICAL PARAMETERS					SPE	CIFIED	RESULT	
1	One-piece vii	nyl upholstered top wit	h high-der	ity urethane foam Spe		Spec	cify		
2	Table is equipped with 2 front end drawers and 3 right side drawers					Spec	Specify		
3	Electro-mechanical actuator system or full electric system					Spec	Specify		
4	Electrical Treatment table to be used for Ob/Gyn patients .					Yes	Yes		
5	Foot or Hand control .					Yes	Yes		
6	Power up, down, auto return, auto stop, foot extension, pelvic tilt .					Yes	Yes		
7	Including drain pan, hidden stirrups, drawers, Ob/Gyn accessories .					Yes	Yes		
8	Non-skid surface, comfortable built-in pillow .					Yes	Yes		
9	Min height .					Spec	Specify		
10	Max height .					Spec	Specify		
11	Max patient weight > 225 kg					Yes	Yes		
12	Table width .					Spec	Specify		
13	Storage Facility to store paper rolls, instruments and supplies .					Yes	Yes		
14	Preferably with removable upholstered top for easy cleaning or repair .					Yes	Yes		
15	Preferably with infinitely adjustable stirrups, to lock in four lateral positions .					Yes	Yes		
16	Preferably with knee crutches an option .					Spec	Specify		
17	Examination light source integrated with the table .					Spec	Specify		
18	Plastic irrigat	ion pan extends and re	etracts fror	n leg rest		Yes			
19	Pelvic tilt with	n easy-to-use dual side	controls			Yes			
20	All Table Acc	essories for OB/GYN	cases		Yes				
21	Power Supply	y 220 VAC / 60 Hz .				Yes			
22	CE or FDA .					Yes			

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.

يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

<u>02/Aug/2016 11:34 AM</u> Page 1 of 1