

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010106	GROUP #	PHC201
ITEM DESCRIPTION	WHEELCHAIR - ADULT	اسم البند	كرسي متحرك - بالغين

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	The Main structure of the unit shall be from a heavy duty stainless steel or chrome plated steel	Yes	
2	The unit shall have a Two large rear wheels inflatable (complete with manual air pump).	Yes	
3	The unit shall have a Two small front (steering) wheels made of solid rubber (no air).	Yes	
4	The unit shall have a Rear pocket	Yes	
5	The sides of the unit must be Removable	Yes	
6	The unit shall be included with foot rests.	Yes	
7	The Foot rests shall be adjustable and can be opened wide to allow patient to stand up.	Yes	
8	The Seat of made from heavy duty material or leather that is easy to clean	Yes	
9	The Chair must be completely foldable.	Yes	
10	The seat size shall be 18 width ,18depth WEIGHT	Yes	
11	The weight capacity shall be 200 kg or more	Yes	
12	Itemized price for any options and accessories should be provided	Yes	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها