

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010045	GROUP #	PHC092
ITEM DESCRIPTION	FETAL HEART DOPPLER	اسم البند	جهاز قياس نبض الجنين

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	HAND HELD UNIT	Yes	
1	PROBE	5 - 13 MHZ (LIST ALL PROBES)	
1	DISPLAY	LCD	
1	BATTERY OPERATED	INCLUDED	
1	RS-232 INTERFACE	PREFERRED	
1	HEADPHONE OUTLET	PREFERRED	
1	EMI STATUS	PROTECTED AGAINST EMI	
1	OTHER SPECIFICATIONS		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها