

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010021	GROUP #	PHC050
ITEM DESCRIPTION	CAUTRY UNIT	اسم البند	جهاز كي كهربائي مع الملحقات
MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN	
#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
2	Monopolar and bipolar mode	Yes	
3	Monopolar output	30w	
4	Bipolar electrocoagulation	30w	
5	Pencil with cable	Included	
6	Needle straight electrod	Included	
7	Ball electrode	Included	
8	Blade electrode	Included	
9	Bipolar Forceps with cable	Included	
10	Neutral bad	Included	
11	Trolley	Included	
12	Foot Switch	Specify	
13	220 VAC - 60 HZ	Yes	
14	FDA or CE approved	Yes	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها