

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260050007	GROUP #	PHC235
ITEM DESCRIPTION	X-RAY PROTECTIVE ACCESORIES - FOR PHC	اسم البند	إكسسوارات واقية أشعة - للمراكز الصحية

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Adult Apron Set :	One Set with Holder	
1.1	Material	Flexible Multiple Layer of Lead/Lead Free Rubber Sheet with Light Weight	
1.2	Tear Resistance	Yes	
1.3	Easy Clean and Disinfection	Yes	
1.4	Front Protection	Full Length	
1.5	Back Protection	Half Length	
1.6	Thickness	Front : 0.5 mm Pb , Back : 0.25 mm Pb	
1.7	Common Size, cm	Small , Medium , Large	
1.8	Accessories	thyroid Collar 0.50 mm Pb , Standard.	
1.9	Breast Pocket	Yes	
1.10	Closure	Velcro or Belt	
1.11	Sample should be submitted	Color Upon Delivery	
1.12	Other Specs.	Specify	
2	Children Apron Set :	One Set with Holder	
2.1	Material	Flexible Multiple Layer of Lead/Lead Free Rubber Sheet with Light Weight	
2.2	Tear Resistance	Yes	
2.3	Easy Clean and Disinfection	Yes	
2.4	Front Protection	Full Length	
2.5	Back Protection	Half Length	
2.6	Thickness	Front : 0.5 mm Pb , Back : 0.25 mm Pb	
2.7	Common Size, cm	Small , Medium , Large	
2.8	Closure	Velcro or Belt	
2.9	Sample should be submitted	Color Upon Delivery, Rubber Sheet with Drawing for Children	
2.10	Other Specs.	Specify	
3	Patient pelvis lead protection :	One Set with Holder	
3.1	Material	Flexible Multiple Layer of Lead/Lead Free Rubber Sheet with Light Weight	
3.2	Tear Resistance	Yes	
3.3	Easy Clean and Disinfection	Yes	
3.4	Common Size, cm	Small , Medium , Large	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضته المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
3.5	Sample should be submitted	Color Upon Delivery	
3.6	Other Specs.	Specify	

<p>THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.</p>	<p>يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضة المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها</p>
--	---

