

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260090070	GROUP #	PHC302
ITEM DESCRIPTION	WASHER DISINFECTOR 5DIN - SINGLE DOOR	اسم البند	جهاز غسيل آلات 5 دن - باب مفرد

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	application	for mechanical cleaning and disinfection surgical devices	
2	type	floor standing	
3	DISPLAY	LCD	
4	CHAMBER MATERIAL	High quality stainless steel (AISI 316 Ti/AISI 316 L)	
5	CAPACITY	5 DIN	
6	A0 Value monitor	yes 600-3000 approximately	
7	LOADING AND UNLOADING	yes for 2 racks	
8	water softner	Yes	
9	cycle time	specify	
10	wash action	rotary adjustable .spray arms.top,and botton	
11	cold wash	yes	
12	detergent wash1	yes	
13	detergent wash2	yes	
14	thermal disinfection	Yes according A0 value	
15	PRINTER	Yes	
16	drying	yes	
17	heat source	built in electric	
18	booster tunk	specify	
19	Door	Single doors sliding	
20	erorr indication alarms	Yes	
21	power SUPPLY	220/380volt3-phase 60Hz	
22	ACCESSORIES	anstheisa rack,instrument rack ,utensil rack .lumens rack with dental hand piece adaptor .shoes racks,	
23	REGULATORY COMPLIANCE	CE,FDA.. Machine should comply with ISO 15883	
24	OTHER SPECIFICATION	1)supplier should provide a detergent for 3 monthes consumables according to MOH specs . 2)Vendor is responsible for complete installation and demonstration at site	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضة المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها