

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260090063	GROUP #	PHC283
ITEM DESCRIPTION	DENTAL HANDPIECE - LUBRICANT SYSTEM	اسم البند	جهاز تشحيم قبضات الأسنان

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	APPLICATION	Automatically process to cleans ,lubricates ,and dryer	
2	CONFIGURATION	Bench top	
	Handpieces manufacture brands	include high speed handpieces, low speed handpieces, and air motors, e-type attachments etc.	
3	CONTROLS	Microprocessor, fully automatic	
	The system shall be process	at least four handpieces at the same time	
4	The process	Begins with the removal of all visible debris and running the handpiece (with the bur in place) for 20-30 seconds to expel debris from the chuck, turbine, and ball bearing race. The handpiece is then scrubbed, rinsed, and dried.	
13	POWER SUPPLY	220 V, 60 HZ	
14	ACCESSORIES	Stainless steel basket & Instrument tray	
15	REGULATORY COMPLIANCE	1) CE by approved organization for all of the following directives: -Medical Device Directive. -Electromagnetic Compatibility. -European Pressure Equipment Directive -including identification of certifying companies with their EC-registration number 2) FD	
16	OTHER SPECIFICATION	Vendor is responsible for complete installation and demonstration at site	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضة المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها