

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010064	GROUP #	PHC128
ITEM DESCRIPTION	PATIENT BED - ELECTRICAL	اسم البند	سري مريض كهربائي

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Trendelenburg / reverse Trendelenburg	Yes	
2	Mattress has to be included with the bed	Yes	
3	The motor should be heavy-duty type and protected for over heat	Yes	
4	The mattress should be supported by steel frame with molded edges for safety; the frame should have mattress stops	Yes	
5	Side rail	Yes	
6	Bed length:228cm,Width:110cm, Underbedclearance:13cm. approx	Yes	
7	High/Low positioning. Height in high position:75cm and in low position :45. approx.	Yes	
8	Removable headboard for CPR	Yes	
9	4 – Position IV pole holders	specify	
10	Drainage bag holders and Urinal devices	Yes	
11	Easy maneuverability – 4 castors castor size 13cm. approxi.	Yes	
12	Bed shall hold 220 kg working load approx.	Yes	
13	Bumpers	Yes	
14	Battery back-up	Yes	
15	Four corner central brake and steer.	specify	
16	The unit should have no sharp edges	Yes	
17	All external components should be securely mounted	Yes	
18	The unit should be secure and provide adequate protection against moving and electrically energized parts	Yes	
19	The unit should be well constructed with durable materials to withstand typical abuse and cleaning	Yes	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها