

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260090055	GROUP #	PHC275
ITEM DESCRIPTION	BENCH TOP STEAM STERILIZER	اسم البند	جهاز تعقيم - فوق الطاولة

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	TYPE	Table top, B-Class	
2	BODY CONSTRUCTION	Stainless steel 304L Powder coating housing	
3	CHAMBER MATERIAL	Stainless steel, 316L	
4	APPLICATION	Wrapped or unwrapped solid or hollow items Porous loads	
5	CONTROLS	Microprocessor, fully automatic	
6	Bowie & Dick Test and Vacuum Test	Yes	
7	Special user-programmable cycle	Specify	
8	LOW WATER LEVEL ALARM	Yes	
9	CYCLE FINISH ALARM	Yes	
10	CYCLE MALFUNCTION	Audio and visual	
11	CYCLE STATUS DISPLAY	Yes, LCD, TOUCH SCREEN	
12	CHAMBER CAPACITY	20 - 25 Liters approx.	
13	LOADING	Front	
14	DOOR	Single, hinged	
15	CYCLES	121, 134 flash, programmable	
16	HEATING ELEMENT CAPACITY	3.5 – 5 Kw, approx.	
17	WORKING PRESSURE	3.5 bar approx.	
18	Active – forced air removal Yes	Yes	
19	PRINTER	Yes	
20	Manual or Automatic water fill and continuous direct drain	Specify	
21	Serial port for data transmission	Specify	
22	Efficient – load up to 6 trays or 4 cassettes	Specify	
23	POWER SUPPLY	220 V, 60 HZ	
24	ACCESSORIES	Shelf tray NOT LESS THAN 6	
25	WATER DISTILLER 5 to 10 Liters approx.	INCLUDED	
26	self-rinsing system	Yes	
27	REGULATORY COMPLIANCE 1) CE and FDA approved	Yes	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها