

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260040003	GROUP #	PHC019
ITEM DESCRIPTION	ASPIRATOR - DENTAL	اسم البند	جهاز شفط للأسنان

MODEL OFFERED		MANUFACTURER	COUNTRY OF ORIGIN
#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	ASPIRATOR, WITH INDUCTION MOTORS FOR CONTINUOUS SERVICE WITH OVER HEAT & OVERFLOW PROTECTION.	Yes, 2800 rpm	
1	MOTOR NOT LESS THAN 1KW	Yes	
1	ALUMINUM SUCTION FAN WITH AIR FLOW NOT LESS THAN 950 L/min..	Yes	
1	STATIC DEPRESSION NOT LESS THAN 1300mm WATER.	Yes	
1	AUTOMATIC, UPON LIFT AND HANG UP BY AUTOMATIC HANGER SWITCH FOR SUCTION TIP.	Yes	
1	AUTOMATIC WATER SEPARATOR FOR CONTINUOUS SERVICE WITH INTERNAL CLEANING SYSTEM EVACUATION PUMP	Yes	
1	CONTROL PANEL	Digital Control	
1	DISPLAY	1 - 2 Dental unit	
1	CAPACITY		
1	POWER SUPPLY	220V, max. 15A, 1 Phase	
1	VOLTAGE	60Hz	
1	FREQUENCY	3Pin BT, 1-2 Meter Length	
1	POWER CORD	All standard Provide list offer	
1	ACCESSORIES	ISO, FDA, CE, or other. Specify in offer	
1	REGULATORY COMPLIANCES	EMC,UL, IEC, CSA or other Intl. Standard, (Specify in offer).	
1	SAFETY STANDARD		
1	OTHER SPECIFICATIONS		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها