

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010033	GROUP #	PHC073
ITEM DESCRIPTION	CURTAIN - CUBICAL	اسم البند	ستارة لسرير الكشف معلقة بالسقف ومتحركة

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	inpatient cubical curtain	نعم	
1	inpatient cubical curtain	نعم	
2	fabric		
3	Weight	200 g/sqm Min.	
4	Light fastness	5to 7	
5	Light reflexion	55% Min.	
6	light transmission	30% Min.	
7	Solar reflexion	50% Min.	
8	Solar transmission	30% Min.	
9	Antibacterial Treatment	Bioactive	
10	Cleaning	machine washable up to 60 degrees Celsius Enviromental friendly , 100% Free of PVC , halogene and formaldehyde	
11	Fullness	1 to 1.25 per Meter	
12	Colour	Variety	
13	HIGHT	THE CURTIN FABRIC MUST RISE ABOVE THE FLOOR 20-25 CM AND THE FABRIC HIGHT 240-250 CM .	
14	Curtain tracks		
15	Material	Aluminum , white cromated powder coated SMOTH CURVES	
16	Dimensions	25mm x 30mm Min.	
17	Weight	400 grams per meter Min.	
18	Hocks	Very high quality s/s	
19	other		
20	The vendor is responsible for installation of CUBICAL curtains with all accessories needed		
21	Has permanent antimicrobial effect without additional finishing agent		
22	Agent cannot be washed out , since integrated into fibre		
23	Minimises the risk of transferring bacteria		
24	Antimicrobial effect certified		
25	Permanent Anti-Bacterial Effect	Built-In	
26	Light Fastness		
27	Satisfies highest hygienic demands		
28	Resistance to all methods of disinfection		
29	Easy care , hardwearing , Low- crease and dimentionally stable		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
30	Short wash cycles at low temp. save water , detergent and energy		
31	High colour brilliance and fastness to light		
32	Does not decay or mould		
33	High resistance to abrasion		
34	Flame retardant fibre and filament yarns are free of toxic substances		
35	Meets all important international fire protection standards e.g. DIN 4102 Part 1/B1; OENORM3820/B1, NF P 92501-7/M1;BS5867, Part 2 type C; UNI VF 8456/I ; UNE 23721-27/M1 and EN 1021 Part I 1+2		
36	No after-glow and only very slight development of fumes in the event of fire		
37	Supported by general test certification		
38	warranty		
39	5 years , fabric,track and components installation to the ceiling for the entire project		
40	important	supplier must visit site to provide correct dimension, and submit certificate of visit signed by respective moh /signed by hospital authority with the offer to ensure correct dimension and type of casework will be supplied) upon visiting the site ,suppl	
41	ON-SITE ACTUAL METERS OF THE PROJECT	VENDOR MUST PROVIDE THE ACTUAL EXACT LENGTH OF CURTAIN NEEDED FOR ALL AREA WITH-IN THE HOSPITAL	
42	Price		
43	SUPPLIER MUST PROVIDE THE PRICE AND THE QTY OF EACH PART OF THE CURTIN ACCORDING TO THE DOWN-MENTIONED SPECS.		
44	ONE LENGTH METER OF TRUCK WITH FABRIC (PRICE)		
45	TOTAL QTY OF THE PROJECT (LENGTH METER)		
46	TOTAL PRICE FOR THE PROJECT		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضة المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها