

**TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS**

**التجهيزات الطبية**

**مراكز الرعاية الصحية**

<b>Item Code</b>	4455260010041	<b>GROUP #</b>	PHC088
------------------	---------------	----------------	--------

<b>ITEM DESCRIPTION</b>	EXAM TABLE	<b>اسم البند</b>	طاولة فحص
-------------------------	------------	------------------	-----------

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Used for general patients	Yes	
2	Electrical System	Yes	
3	Adjustable height by using foot switch	Yes	
4	Foot control	Yes	
5	Power up, down, auto return, auto stop, foot extension, pelvic tilt	Yes	
6	power pack	Yes	
7	Including drain pan, hidden stirrups and drawers	Yes	
8	Non-skid surface, comfortable built-in pillow	Yes	
9	Max patient weight > 225 kg	Yes	
10	Storage Facility to store paper rolls, instruments and supplies	Yes	
11	Removable upholstered top for easy cleaning or repair	Yes	
12	Infinitely adjustable stirrups, to lock in four lateral positions	Yes	
13	knee crutches	Option	
14	Examination light source integrated with the table	Specify	
15	Dimensions approximately (L =220 cm, W = 65 cm)	Yes	
16	Minimum Height	45 cm approx.	
17	Maximum Height	94cm approx.	
18	Power supply 220 VAC / 60 Hz	Yes	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها