

**TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS**

**التجهيزات الطبية**

**مراكز الرعاية الصحية**

<b>Item Code</b>	4455260010056	<b>GROUP #</b>	PHC119
<b>ITEM DESCRIPTION</b>	NEBULIZER ULTRASONIC	<b>اسم البند</b>	مرذاذ فوق صوتي

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	NEBULIZER		
2	Warm-up time, min	<1.5	
3	Mist temperature, 0C	33-35 aprox.	
4	Mist output, mL/min	0.2-3.15	
5	Noise level, dbA,		
6	max @ 1 m distance	<40	
7	O2 conc settings, %	40-100 % approx.	
8	HEATING UNIT		
9	Location	Tube	
10	Type	Processor controlled/heater block	
11	Power-on indicator	Yes	
12	Temperature control	Yes	
13	Safety shutoff	Yes	
14	Line voltage, VAC	220-240 V, 60 Hz	
15	RESERVOIR		
16	Refillable	YES	
17	Usable volume, mL	>200	
18	Water ml	>200	
19	Saline (0.45/0.9%)	>150	
20	STERILIZATION METHOD		
21	Nebulizer head and Heating unit	Chemical and autoclaveable	
22	Reservoir	Chemical and autoclaveable	
23	OTHER SPECIFICATIONS	ALL ACCESS. NEEDED TABLE TOP	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

Kingdom of Saudi Arabia Ministry Of Health General Directorate of Equipment Supplies Department Of Specs. & Development	 وزارة الصحة Ministry of Health	المملكة العربية السعودية وزارة الصحة إدارة التجهيزات الطبية إدارة المواصفات والتطوير
--	---	---

# TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

--

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضة المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها