

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260090051	GROUP #	PHC271
------------------	---------------	----------------	--------

ITEM DESCRIPTION	BIOLOGICAL INDICATOR READER	اسم البند	جهاز قراءة بيولوجي
-------------------------	-----------------------------	-----------	--------------------

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	TYPE	incubator	
2	incubator	Complies with (Rapid readout Biological indicator) item number (124990044) required from the Department of Medical Supply	
3	printer	yes built in	
4	result	yes for positive ,negative and cancelled sample	
5	authorized standard	ISO 13485	
6	temperature	dual temp.37 to 60 approximately	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها