

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260050005	GROUP #	PHC083
ITEM DESCRIPTION	DRY PRINTER	اسم البند	طابعة أفلام أشعة

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Tabletop type	Yes	
2	Dry Printer	Yes	
3	pixels per inch (Resolution)	320 DPI	
4	Speed of at least 40 films / hour for 35 X 43 cm size	Yes	
5	Automatic Density Correction	Yes	
6	Film sizes can be printed, cm	(20 x 25), (25 x 30), (35 x 35), (35 x43), (28 x 35)	
7	Gray Scale resolution of 14 bits	Yes	
8	Daylight film loading	Yes	
9	Network Connectivity	Yes	
10	Power Requirement	220-240 V / 60 Hz	
11	Other Specifications	Specify	
12	Warranty	5 years full warranty Unlimited including All Parts	
13	Maintenance and Service Contract Value with Spare Parts for the First Year after Warranty	Specify Value	
14	Maintenance and Service Contract Value with Spare Parts for the Second Year after Warrant	Specify Value	
15	Maintenance and Service Contract Value with Spare Parts for the Third Year after Warranty	Specify Value	
16	Maintenance and Service Contract Value with Spare Parts for the Fourth Year after Warranty	Specify Value	
17	Maintenance and Service Contract Value with Spare Parts for the Fifth Year after Warranty	Specify Value	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها