Kingdom of Saudi Arabia Ministry Of Health General Directorate of Equipment Supplies Department Of Specs. & Development



المملكة العربية السعودية وزارة الصحة إدارة التجهيزات الطبية إدارة المواصفات والتطوير

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

Item Code					
	MODEL OFFERED	MANUFACTURER		COUNTRY OF	ORIGIN
	TECHNICAL PARAMETERS		SPECIFI	ED	RESULT
	TYPE		Floor star	Floor standing	
	CONFIGURATION		Storage of under safe	Storage of pharmaceuticals under safe, controlled conditions	
	CAPACITY IN LITERS		700 L Ap	700 L Approximatlly	
	SHELVES			4 or 5 heavy gauge zinc plated /powder coated	
	DOORS				
	DOOR TYPE		Glass do	Glass door	
	DOOR SEALING SYSTEM		Yes, para	Yes, para-magnetic	
	DOOR LOCK		Yes	Yes	
	TEMPERATURE CONTROL		2 degree degree C	Centigrate TO 10 Centigrate	
	INTERIOR		Stainless	Stainless steel	
	EXTERIOR		Powder coated enemal paint		
	TEMP-ALARM . And Door Open Alarm			Yes, High & Low / and Yes for Door alarm	
	TEMPERATURE INDICATOR		Yes, Digi	Yes, Digital display	
	TEMP. CHART RECORDER		Yes	Yes	
	POWER FAILURE ALARM, AUDIO & VISUAL		Yes	Yes	
	RAPID TEMP. RECOVERY AFTER DOO	RAPID TEMP. RECOVERY AFTER DOOR OPENINGS Yes			
	INSULATION		CFC free	CFC free	
	INSIDE LIGHT		Yes		
	CASTERS		Yes 4, (tv	wo lockable)	
	POWER SUPPLY		220 V, 60) Hz	
	OTHER SPECIFICATION		FDA OR	CE	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.

لِلنَرْم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتَعبُّــة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

<u>02/Aug/2016 11:43 AM</u> Page 1 of 1