

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260060036	GROUP #	PHC118
-----------	---------------	---------	--------

ITEM DESCRIPTION	MIXER VORTEX	اسم البند	خلاط لعينات الدم
------------------	--------------	-----------	------------------

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Stirring action at atouch with test tubes	Yes	
2	Stirring pad (ribrating foam pad)	Yes	
3	Mounted on rubber feet	Yes	
4	Varible speed settings for gental and through shaking action	Yes	
5	Contious Mode	Yes	
6	Touch Mode	Yes	
7	SPEED	100 ? 3000 RPM	
8	Intermittent mixing	Yes	
9	Power Supply		
10	Voltage	220Volts	
11	Frequency	60Hz	
12	OTHER SPECIFICATIONS		

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED "COLUMN THEN STAMP AND SIGN THE SHEET.

يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

