

TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS

التجهيزات الطبية

مراكز الرعاية الصحية

Item Code	4455260010039	GROUP #	PHC086
ITEM DESCRIPTION	EXAM LIGHT - CEILING MOUNTED	اسم البند	كشاف عمليات سقفى

MODEL OFFERED	MANUFACTURER	COUNTRY OF ORIGIN

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
1	Type and number of light heads	Ceiling mounted/single light head	
2	Number of bulbs	2	
3	Color TEMP, K	3,500 K	
4	Color Rendering		
5	Index	APPROX 90	
6	Field size, cm		
7	Diameter	30 cm	
8	Depth	100cm	
9	Focal length	Auto focus	
10	Illumination level		
11	Intensity	110,000 lux	
12	Light patch diameter	25 cm	
13	Controls		
14	Dimmer	Variable	
15	Rotation	360?	
16	Vertical adjustment	Yes	
17	Filter material	Dichroic mirrors	
18	Sterilizable handle	Yes	
19	Supply Voltage VAC	220-230	
20	Bulb type	Halogen	
21	Volts	24 V	
22	Easy to replace	Yes	
23	Working distance	200 cm	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتعبئة عرضه المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها

**TECHNICAL & PERFORMANCE SPECIFICATION FOR MEDICAL EQUIPMENTS**

#	TECHNICAL PARAMETERS	SPECIFIED	RESULT
24	Other specification	20 spare bulbs	

THE BIDDER MUST COMPLY WITH ALL ABOVE SPECS. AND MUST FILL HIS OFFER IN THE "OFFERED" COLUMN THEN STAMP AND SIGN THE SHEET.



يلتزم مقدم العطاء بكافة المواصفات الفنية الموضحة أعلاه كما يلتزم بتقديم عرضته المقدم على أوراق المواصفة بالخانة الموضحة مع ختمها وتوقيعها