

Iontophoresis Drug Delivery Integrated With Physiological Detection and Preventive Care

Executive summary

Iontophoresis is re-emerging as a serious wearable-drug-delivery platform because recent work has moved beyond legacy benchtop stimulators toward flexible patches, integrated power sources, hydrogel-electrode composites, porous or solid microneedle hybrids, and on-skin sensing architectures. Over the last five years, the strongest advances have come from three converging streams: compact iontophoretic delivery patches, wearable biosensors that can detect actionable physiology from sweat or interstitial fluid, and hybrid systems that merge transdermal transport with on-body electronics. Reviews from 2021 to 2024 consistently frame the field as transitioning from “enhanced transport” to “intelligent transdermal systems,” while reverse iontophoresis and wearable patches have benefited especially from flexible electronics and patch-level integration. ¹

The recent device literature is materially better than the field was even a few years ago. Examples include a cellphone-controlled constant-current iontophoretic patch with Bluetooth communication, a pressure-ulcer proof-of-concept that uses the same flexible electrode pair for impedance sensing and drug delivery, a built-in Mg-battery patch using viologen hydrogels as both drug reservoir and cathode, a dual-responsive Zn-battery hydrogel patch for atopic dermatitis, a paper-battery microneedle-iontophoresis system for hypertrophic scars, and a 2026 eco-degradable self-powered patch whose electrochromic gauge visually tracks cumulative delivered charge. Together, these papers show that dose control, portable power, thin flexible form factors, and user-facing dose indication are no longer speculative. ²

What is still missing is equally important. Fully therapeutic, human, closed-loop iontophoretic drug-delivery systems that use real-time physiological sensing to autonomously modulate drug dose remain rare. The literature is richer in: remote-controlled iontophoresis, self-indicating dose-tracing patches, monitoring-only epidermal patches, wound-monitoring electrotherapy systems, and hybrid microneedle or electroosmotic systems. Based on the recent examples and reviews surveyed here, this means the near-term opportunity is not a fully autonomous “artificial pancreas by iontophoresis,” but a safer class of semi-closed-loop or supervised closed-loop patches for local anti-inflammatory therapy, early wound/pressure-ulcer prevention, scar therapy, or sensor-guided rescue dosing. That conclusion is an inference from the current evidence base rather than a direct statement from any single source. ³

Because the target drug, patient population, body site, intended claim, and care setting were not specified, this report treats those items as open-ended design variables. The most practical project path is to choose an initial indication where local delivery, modest dose, and measurable skin physiology all align: dermatitis flare suppression, early pressure-ulcer prevention, hypertrophic scar control, or sensor-guided local analgesia. Those use cases match the current maturity of iontophoretic materials, electronics, and regulatory precedent better than systemic hormone replacement or fully autonomous metabolic therapy. ⁴

Design space and current state of the art

Iontophoresis enhances transdermal transport by applying low direct current across the skin, leveraging electromigration and electroosmosis. Its appeal is straightforward: non-invasive administration, temporal controllability, compatibility with wearable form factors, and the ability to improve flux for charged or polar molecules while avoiding needles in many use cases. Its limitations are equally foundational: the stratum corneum remains a strong electrical and mass-transfer barrier; conventional iontophoresis strongly favors relatively small ionized molecules; and current density, pH drift, heat, and contact quality directly affect both delivery and irritation risk. ⁵

The most concrete recent improvement has been patch-level integration. In 2023, Mori and colleagues demonstrated a remote-controllable dosage-management system in which a cellphone acted as the transmitter, a patient-worn receiver/dosing unit served as the patch controller, Bluetooth linked the two, and the therapeutic node was a constant-current iontophoresis device tested on excised pig skin. That study matters less for its specific pharmacology than for showing a viable electronics architecture for supervised dosing and telemedicine-adjacent control. ⁶

A second important direction is preventive-care integration. Chaudon and co-workers built an all-in-one proof-of-concept for pressure ulcers that used impedance spectroscopy for early detection and iontophoretic delivery of tolazoline for localized treatment. The same flexible electrode pair handled both sensing and therapy, and the compact current source increased penetration by up to eightfold at 250 μA versus passive diffusion; the system achieved up to 31.65 $\mu\text{g}/\text{cm}^2$ tolazoline in skin during a 30-minute treatment. This is one of the clearest recent demonstrations of body-physiology detection directly coupled to iontophoretic intervention. ⁷

A third direction is self-powered or battery-integrated delivery. In 2023, a Nature Communications study reported a simplified wearable patch with a built-in Mg battery and viologen-based hydrogels that functioned as both integrated drug reservoir and cathode material, thereby reducing interface impedance and device complexity. With dexamethasone sodium phosphate loaded into the hydrogel, the patch showed a 3.2-fold permeation increase relative to passive hydrogel control after three hours, with roughly threefold more drug deposited in deeper epidermal/dermal layers. In a psoriasis mouse model, the same platform restored lesional skin morphology toward normal after repeated treatment. ⁸

That theme continued in 2025 and 2026. A 2025 atopic-dermatitis study used a sulfonated PEDOT polyelectrolyte hydrogel as both reservoir and cathode in a Zn-battery iontophoresis patch; the dual-responsive hydrogel combined endogenous thermal response and exogenous electrical stimulation, reached 68.4% release efficiency at 37 °C, and delivered 12.2 $\mu\text{g}/\text{cm}^2$ dexamethasone phosphate over three hours in hairless mouse skin. In 2026, an eco-degradable self-powered iontophoretic patch synchronized drug delivery with electrochromic dose indication, using an Mg–MoO₃ galvanic cell and a WO₃-based electrochromic gauge so that cumulative charge became a visible proxy for delivered dose, without external sensors or conventional control electronics. ⁹

Hybrid iontophoresis-microneedle systems are now one of the most dynamic subfields because they directly tackle the main physical limitation of classical iontophoresis: skin barrier resistance and the poor transport of larger molecules. A 2021 porous-microneedle Nature Communications paper showed that porous microneedles with charged hydrogels simultaneously lowered transdermal resistance, enabled transport of larger molecules, and generated electroosmotic flow. The authors demonstrated delivery of 10

kDa dextran and extraction of glucose through pig skin, and even powered the system with a fructose/O₂ enzymatic biobattery. In 2023, a wearable vaccine macromolecule platform used flexible-PI Ag/AgCl electrodes, solid microneedles, hydrogel blocks, an impermeable adhesive gasket, and a constant-current module to outperform intramuscular administration in vivo. In 2025, a paper-battery microneedle-iontophoresis patch for hypertrophic scars actively delivered 90.19% of triamcinolone acetonide into scar tissue and reduced TGF- β 1 and collagen I expression. ¹⁰

The adjacent closed-loop literature is also informative even when delivery is not purely iontophoretic. A 2024 diabetes-management patch combined a graphene-composite ink-printed glucose sensor on hollow microneedles, a PEG-functionalized electroosmotic micropump, and a PCB controller in a wearable system that achieved continuous glucose control in diabetic rats. This is not pure iontophoresis, but it shows what the control stack around an intelligent transdermal platform can look like once sensing, actuation, and embedded electronics are fully integrated. ¹¹

Sensors, control, electronics, and firmware

For physiology-aware or preventive-care iontophoresis, the sensor menu is already broad enough to support practical prototypes. Recent epidermal and microneedle patches have shown that blood-pressure and heart-rate ultrasound, sweat chemistry, interstitial-glucose sensing, impedance sensing, skin temperature, and deep-learning-based transdermal biomarker detection can all be worn in patch form. The challenge is not whether sensing is possible; it is whether the sensed signal is sufficiently causal, stable, and actionable to justify modulating dose in real time. ¹²

The strongest sensor classes for near-term iontophoretic control are skin impedance, local temperature, sweat chemistry, and selected interstitial-fluid analytes. Skin impedance is especially attractive because it is directly relevant to both safety and dose-transfer efficiency: it can indicate skin contact quality, hydration, barrier state, wound change, and resistance drift during dosing. That direct coupling is why the pressure-ulcer patch reused the same electrodes for impedance measurement and iontophoresis, and why recent wound systems use impedance sensing to quantify healing state. Temperature is similarly useful because it affects both skin physiology and analyte behavior, and FDA specifically flags heat as a safety/performance issue for transdermal systems. ¹³

More ambitious sensing options are feasible but indication-dependent. The 2021 all-in-one epidermal patch combined ultrasound haemodynamics with glucose-in-ISF and sweat alcohol, caffeine, and lactate sensing; a 2024 sweat patch used silicon ISFETs plus temperature sensing to monitor pH, sodium, and potassium during 85 minutes of exercise; a 2024 polymer microneedle patch continuously measured interstitial glucose through impedance over up to seven days; and a 2023 wrist-worn transdermal troponin-I sensor used a CNN-RNN model to predict elevated hs-cTnI with internal/external AUROCs of 0.90 and 0.92. These are not, by themselves, iontophoretic controllers, but they validate the sensing modalities that a closed-loop iontophoretic system can be built around. ¹⁴

Candidate physiological sensors for iontophoresis-integrated systems


Sensor modality	What it measures	Why it matters for iontophoresis or preventive care	Integration maturity with patch systems	Main limitations
Skin impedance / bioimpedance	Contact quality, hydration, barrier integrity, local tissue change	Most directly useful for safety interlocks, adaptive current control, and early wound/pressure-ulcer detection; already demonstrated in combined sensing-treatment patches. ¹⁵	High for patch integration	Signal can drift with sweat, motion, pressure, and electrode polarization
Skin temperature	Local thermal state	Useful for dose compensation, irritation detection, and analyte correction; also relevant because heat can alter transdermal performance. ¹⁶	High	Temperature alone is rarely a sufficient therapeutic trigger
Sweat pH / Na / K	Sweat electrolyte balance and local sweat chemistry	Best for hydration, training, skin condition, and contextual dosing rather than direct drug control; mature multiplexed patch examples exist. ¹⁷	Moderate to high	Requires stable sweat generation and compensation for flow and temperature
Interstitial glucose	Metabolic state	High-value for diabetic or steroid-linked glycaemic monitoring; microneedle and reverse-iontophoresis approaches are advancing. ¹⁸	Moderate	Pure iontophoretic closed-loop glucose-to-drug control is still immature
ECG / PPG / haemodynamics	Heart rhythm, HR, HRV, BP proxies	More useful for preventive monitoring and contraindication handling than for direct iontophoretic titration; multimodal chest/epidermal patches are feasible. ¹⁹	High	Ambiguous causal link to most transdermal drugs unless indication is cardiovascular


Sensor modality	What it measures	Why it matters for iontophoresis or preventive care	Integration maturity with patch systems	Main limitations
Transdermal troponin or other ML-derived biomarker signals	Acute biomarker risk state	Potentially valuable for preventive triage and event-triggered treatment logic, but still early and algorithm-heavy. ²⁰	Early	Clinical validation, explainability, and regulatory burden are substantial

The electronics stack in recent papers converges on a small set of patterns. Recent systems use either a controlled current source plus a battery/boost stage, or a galvanic cell that itself becomes the power source. The pressure-ulcer device used a boost converter feeding a programmable current source; the 2023 vaccine microneedle system used a flexible PCB and an LM334-based constant-current output module; the cellphone-controlled patch separated a phone-side transmitter from a patient-side receiver and current-delivery node connected by Bluetooth; and the 2024 diabetes patch used a PCB for intelligent control of sensing and actuation. This is enough evidence to recommend an architecture built around constant-current delivery, continuous skin-impedance monitoring, and watchdog-controlled fail-safe shutoff. ²¹

The literature is also beginning to clarify firmware strategy. When the physiological trigger is simple and local, a deterministic state machine is usually preferable to a black-box model: it is easier to verify, safer to constrain, and easier to regulate. Where the physiological signal is indirect or noisy, recent patch work suggests adding an estimation layer rather than directly dosing off raw measurements. The 2026 electrochromic patch demonstrates that cumulative charge can serve as a physically grounded dose proxy, while the 2023 troponin paper shows that ML can add value when the sensed signal is indirect and multimodal. A well-designed iontophoretic firmware stack should therefore separate sensing, estimation, control, and safety arbitration into distinct modules. ²²

The hardware block diagram below reflects architectures reported in recent wearable iontophoresis, wound-monitoring, vaccine-microneedle, and remote-controlled patch systems. ²¹






A firmware and control partition for a project-grade prototype can remain modest and still be rigorous. 23

```

[Sensing tasks]
- skin impedance
- temperature
- optional sweat / ISF analyte
- current / voltage logging
  |
  v
[Signal conditioning]
- filtering
- artifact rejection
- calibration / temp compensation
  |
  v
[State estimation]
- skin-contact quality
- barrier/hydration state
- estimated delivered charge / dose
- risk flags
  |
  v
[Controller]
- open-loop recipe
- threshold logic or PID/MPC overlay
- dose caps and refractory periods
  |
  v
[Safety arbiter]
- max current density
- max cumulative charge
- pH / temp / impedance alarms
- power fault and motion fault handling
  |
  
```

The following flowchart is the most defensible closed-loop logic for near-term development: physiology informs *permission and modulation* of iontophoresis rather than unconstrained autonomous dosing. That framing matches the strongest current evidence. 24



In practice, three algorithmic tiers are enough for most first-generation systems. Tier one is deterministic: threshold logic and timed current pulses with hard safety caps. Tier two adds adaptive control: impedance-compensated current scheduling, temperature compensation, and cumulative-charge dose estimation. Tier three is inferential: ML or hybrid estimators when the relationship between measured physiology and therapeutic need is indirect, as in transdermal troponin prediction or multimodal cardiometabolic monitoring. For most iontophoretic preventive-care prototypes, Tier one or two is the right starting point. 25

Materials, suitable drugs, and device architectures

The materials trend is clear: the field is moving away from a strict separation between “electrode,” “drug reservoir,” and “patch substrate.” Recent work increasingly merges those roles through conductive hydrogels, printable composites, and flexible layered stacks. Viologen hydrogels in the Mg-battery patch acted as both reservoir and cathode; sulfonated PEDOT polyelectrolyte hydrogel in the atopic-dermatitis patch also served as both reservoir and cathode; the vaccine microneedle system used conductive

polyacrylamide/chitosan hydrogels for vaccine storage and iontophoretic conduction; and the 2026 eco-degradable patch used PBAT substrates, Mg–MoO₃ galvanic electrodes, agarose gels, and a WO₃ electrochromic gauge to make the entire system thin, soft, and intermittently disposable. ²⁶

Hybrid microneedles change the drug-selection problem. Classical iontophoresis still most cleanly fits small ionized molecules, especially for local therapy. But microneedle-assisted pathways, porous-microneedle electroosmosis, and paper-battery or constant-current microneedle platforms now make larger macromolecules and poorly permeable targets more realistic. That does not mean every macromolecule is now a good iontophoresis candidate; it means the practical frontier is shifting from “small ionic drugs only” toward “small ionic drugs first, macromolecules with barrier-bypass hybrids second.” ²⁷

Drug classes that are most suitable for iontophoresis or iontophoresis-hybrid delivery

Drug or payload class	Example	Why it fits	Recent or relevant evidence
Small cationic local anesthetics	Lidocaine HCl	Strong classical fit for iontophoresis; fast local action; good early prototype drug because efficacy can be measured in vitro and ex vivo	Human-skin work showed therapeutic permeation within minutes at 0.5 mA/cm ² . ²⁸
Small anionic anti-inflammatory steroids	Dexamethasone sodium phosphate	Good for localized inflammatory disease, scar, and dermatology models; directly demonstrated in recent battery-powered patches	Mg-battery patch and Zn-battery AD patch both used dexamethasone phosphate with successful ex vivo/in vivo performance. ²⁹
Vasodilators for local preventive care	Tolazoline	Strong use case for early pressure-ulcer or ischemic-risk rescue because need is local, not systemic	Pressure-ulcer patch selected tolazoline because it is small and positively charged around pH 5–7. ³⁰
Sweat-induction agents for sensing-enabled systems	Pilocarpine	Useful when the primary therapeutic objective is paired with sweat-based monitoring or quality assurance	Integrated sensing literature uses iontophoresis-assisted sweat induction as part of sensing workflows. ³¹
Macromolecular vaccines	Protein antigens	Poor fit for classical iontophoresis alone, but increasingly plausible with microneedle-iontophoresis hybrids	The 2023 vaccine MN patch improved active transdermal immunization and used constant-current active delivery. ³²

Drug or payload class	Example	Why it fits	Recent or relevant evidence
Intralesional anti-scar steroids	Triamcinolone acetonide	High-value hybrid use case where pain reduction and spatially improved delivery matter more than fully needle-free operation	The 2025 paper-battery microneedle-iontophoresis patch actively delivered 90.19% drug into hypertrophic scar tissue. ³³

A practical selection rule follows from the evidence. If you want a first iontophoresis-only prototype, pick a small ionized local drug such as dexamethasone phosphate, lidocaine HCl, or tolazoline. If you want a high-value translational path and can tolerate more complexity, use a hybrid microneedle-iontophoresis platform for scar, vaccine, or difficult-barrier indications. ³⁴

Electrode and interface material comparison

Electrode or interface material	Typical role	Advantages	Drawbacks	Representative recent evidence
Ag/AgCl on flexible PI	Conventional constant-current iontophoresis electrodes	Stable electrochemistry, familiar fabrication, flexible substrate compatibility	Still requires separate reservoir/interface management	Used in the wearable vaccine MN patch with flexible PI and constant-current output. ³²
Viologen-based conductive hydrogel	Integrated cathode + drug reservoir	Reduces interface impedance, soft contact, merges electrode and depot	Chemistry is more specialized; formulation stability and manufacturability need work	Central to the 2023 Mg-battery dexamethasone patch. ³⁵
Sulfonated PEDOT polyelectrolyte hydrogel	Dual-responsive reservoir + cathode	Thermal plus electrical responsiveness, conductive, hydrogel-friendly for inflammatory skin disease	Requires indication-specific chemistry validation and long-term stability work	Used in the 2025 atopic-dermatitis Zn-battery patch. ³⁶

Electrode or interface material	Typical role	Advantages	Drawbacks	Representative recent evidence
Mg–MoO ₃ galvanic electrode pair	Embedded self-powered iontophoretic source	Eliminates external battery pack, patch-level simplicity, eco-degradable system potential	pH control and discharge stability must be managed	Used in the 2026 self-indicating eco-degradable patch. ³⁷
Porous polymer microneedles plus charged hydrogel	Barrier-bypass interface and electroosmotic conduit	Lowers resistance, supports larger molecules, enables extraction as well as delivery	More complex fabrication and mechanical QC	Demonstrated for larger-molecule delivery and glucose extraction in 2021. ³⁸
Paper battery + MN stack	Self-powered hybrid delivery	Extremely compact, low-profile, self-management-friendly	Disposable power source introduces lot-to-lot consistency questions	Used in the 2025 hypertrophic-scar PBIMNP. ³³

For a first project, there is no reason to start with exotic materials unless the indication demands it. A commercially available Ag/AgCl flexible electrode plus a hydrogel reservoir and a well-characterized constant-current driver is the fastest route to a testable system. Materials such as viologen hydrogels, PEDOT polyelectrolytes, electrochromic gauges, or paper batteries become valuable once you are optimizing integration, disposability, or user interaction rather than merely proving delivery. ³⁹

Safety, regulation, and patents

The core safety issues are electrical, chemical, thermal, and skin-mechanical. Current density matters, but so do treatment duration, electrode area, local pH, buffer chemistry, heat exposure, and barrier state. Older but still relevant histopathology work showed that 0.5 mA/cm² methotrexate iontophoresis produced more persistent skin disruption than 0.2 mA/cm², suggesting that “higher current is better” is a poor design assumption. By contrast, newer self-powered patches often operate in lower current-density regimes such as 50–200 µA/cm² and explicitly manage pH, which is a better starting point for safe wearables. ⁴⁰

pH control is not a secondary detail. The 2026 self-powered patch showed that without buffering, Mg chemistry can alkalinize the local environment; using pH 5 citrate buffer kept the operating environment near neutral and improved skin compatibility. That result is directly relevant to any galvanic or battery-integrated iontophoretic design. ⁴¹

Drug-product quality and patch reliability are also regulatory safety issues. FDA's transdermal/topical guidance emphasizes product design, manufacturing controls, finished-product control, adhesion failure, and the impact of applied heat on delivery. For a physiology-aware wearable patch, that translates into specific engineering obligations: characterize dose versus cumulative charge, quantify impedance drift,

validate adhesive wear under sweat and movement, and prove that moderate heating from the environment, skin, or electronics does not cause unintended overdose. ⁴²

On the device side, FDA's current product-classification database lists iontophoresis devices under 21 CFR 890.5525 as Class II devices with 510(k) pathways for both "specific uses" and "other uses," product codes KTB and EGJ. That is useful precedent for stand-alone iontophoresis hardware, but it does **not** automatically de-risk a drug-device combination patch or a novel closed-loop biomarker-treatment platform. Once the primary mode of action is a drug, or once the sensing/algorithm materially affects treatment decisions, the regulatory logic shifts toward combination-product review and potentially more extensive software and human-factors evidence. ⁴³

In the United States, the Office of Combination Products determines the lead center based on primary mode of action. FDA's combination-product materials make clear that drug-device products can be reviewed through different center pathways depending on whether the primary therapeutic effect is attributable mainly to the drug or to the device. In Europe, EMA states that if the medicinal product provides the principal mode of action, transdermal patches are generally regulated as medicinal products, with Article 117 implications for the device constituent. ⁴⁴

For software, usability, and risk management, FDA-recognized standards provide a practical scaffold even if your prototype is still pre-submission. IEC 62304 is recognized for medical-device software life-cycle processes, IEC 62366-1 for usability engineering, IEC 60601-1 for basic electrical safety and essential performance, and ISO 14971 for device risk management. If your patch includes wireless connectivity or algorithmic adaptation, these standards stop being optional good practice and start becoming part of the evidence package you should build from day one. ⁴⁵

Regulatory pathway comparison

Product concept	Likely U.S. pathway	Likely EU pathway	Key practical implication
Stand-alone iontophoresis hardware with no drug constituent	Class II, 510(k), 21 CFR 890.5525 if substantially equivalent. ⁴⁶	CE-marked medical device under MDR, classification depends on intended use and invasiveness	Best pathway for proving hardware, safety, and current control before combining with a drug
Drug-loaded wearable iontophoretic patch where drug provides primary therapeutic effect	Combination product; CDER-led NDA/ANDA/ 505(b)(2) logic is likely, with OCP assignment if uncertain. ⁴⁷	Medicinal product with device constituent; Article 117 notified-body input may apply. ⁴⁸	Chemistry, adhesion, dose uniformity, and heat/patch performance become central review topics
Sensor-only preventive patch	Device pathway depends on intended use; if diagnostic and novel, De Novo may be needed	CE-marked medical device / IVD logic depends on claim	Regulatory burden depends heavily on whether you claim diagnosis or only monitoring

Product concept	Likely U.S. pathway	Likely EU pathway	Key practical implication
Sensor-guided iontophoresis with therapy-modulating algorithm	Combination product plus software evidence; may need more rigorous software, usability, and clinical validation	Medicinal product with device/software constituent, plus MDR-aligned device evidence	The algorithm becomes part of the regulated therapy, not merely an app feature

Relevant patents and patent families

Patent or family	Published / status	What it covers	Why it is relevant
EP0702579B1 / US5458569A, Wearable iontophoresis system	Foundational, granted/ expired. ⁴⁹	Early wearable iontophoretic platform architecture	Important prior art for basic wearable patch concepts
WO1998013100A1, Iontophoretic drug delivery system, including unit for dispensing patches	Foundational PCT. ⁵⁰	Controller-plus-patch system architecture	Relevant to system-level control, consumables, and patch logistics
US9072883B2 / WO2009125960, Battery-integrated iontophoresis patch	Granted patent family. ⁵¹	Patch-integrated battery construction	Highly relevant to today's compact self-powered patch trend
TW201825131A / TWI629076B, Wearable injection device integrating iontophoresis and microneedle	Granted in Taiwan. ⁵²	Hybrid microneedle plus direct-current wearable structure	Directly relevant to MN-iontophoresis convergence
CN116429854B, Method for improving glucose monitoring accuracy in reverse iontophoresis technology	Granted in China. ⁵³	Accuracy improvement for reverse-iontophoresis glucose monitoring	Relevant to sensing-enabled closed-loop or supervisory architectures
WO2024259501A1, Non-invasive blood glucose meter using reverse iontophoresis	Recent pending PCT. ⁵⁴	Adhesive reverse-iontophoresis glucose sensor concept	Signals continued IP activity around non-invasive sensing inputs for future closed loops

Patent or family	Published / status	What it covers	Why it is relevant
US20190209828A1, Iontophoretic device comprising a movement sensor	Published, later abandoned. ⁵⁵	Motion-sensor-informed iontophoretic device	Not strong freedom-to-operate evidence, but conceptually relevant for condition-aware dosing

These patents are useful as a technology map, not as a formal freedom-to-operate opinion. Several are foundational or adjacent rather than exact matches to modern closed-loop medical wearables, and a real program would still need a dedicated patent attorney review before commercialization. ⁵⁶

Project-grade concepts and experimental roadmap

The table below proposes five distinct project concepts ranging from low to high complexity. Budget figures are rough author estimates in U.S. dollars, assuming access to a standard electronics bench and basic analytical support; they exclude formal GLP toxicology, full animal studies, or clinical costs.

Concept	Complexity	Objective	Core components	Validation protocol	Main risks and mitigation	Timeline	Bud
Impedance-aware anti-inflammatory iontophoresis patch	Low	Build a safe, current-controlled flexible patch for local dexamethasone-phosphate delivery while logging skin impedance and temperature	Ag/AgCl flexible electrodes, hydrogel drug reservoir, constant-current driver, MCU with ADC, impedance-measurement front end, NTC/RTD temp sensor, BLE logging; closest precedents are the remote-control patch and vaccine MN current-control hardware. ⁵⁷	Bench: current accuracy into resistor/skin phantoms; Ex vivo: Franz-cell delivery into porcine skin versus passive control; Safety: pH drift, TEWL, erythema scoring surrogate, contact loss detection	Under-delivery due to contact drift; mitigate with impedance threshold and dose-by-charge accounting. Skin irritation; mitigate by starting below 0.2 mA/cm ² equivalent exposure and using buffered hydrogels. ⁵⁸	3–4 months	\\$4k \$10k

Concept	Complexity	Objective	Core components	Validation protocol	Main risks and mitigation	Timeline	Bud
Pressure-ulcer sentinel patch with rescue iontophoresis	Low-to-medium	Detect early tissue-risk change by impedance and trigger supervised tolazoline iontophoresis for local perfusion support	Shared sensing/delivery electrodes, compact programmable current source, boost converter or small galvanic source, bioimpedance module, pressure/contact logging, adhesive optimized for sacrum/heel use; directly inspired by the 2022 pressure-ulcer proof-of-concept. ⁵⁹	Tissue phantom + ex vivo skin under compressive loading; benchtop model relating impedance change to compression injury; porcine-skin delivery study of tolazoline; end point: sensitivity of risk detection and delivered $\mu\text{g}/\text{cm}^2$ under triggered dosing	False alarms from sweat/motion; mitigate with repeated measurements and confidence scoring. Clinical plausibility risk; mitigate by framing first build as prevention-assist, not autonomous therapy.	4–6 months	\\$10 \\$25


Concept	Complexity	Objective	Core components	Validation protocol	Main risks and mitigation	Timeline	Bud
Dermatitis flare patch with sweat-chemistry and temperature fusion	Medium	Combine local sensing of temperature plus sweat pH/Na/K with dexamethasone-phosphate iontophoresis for semi-closed-loop dermatitis management	Conductive hydrogel reservoir, flexible current source or Zn/Mg battery variant, sweat microfluidics, ISFET or electrochemical ion sensors, temperature sensor, simple on-patch dose indicator; leverages AD patch, sweat patch, and self-indicating patch ideas. 60	Bench: analyte calibration under variable sweat rates and temperature; Ex vivo: delivery versus cumulative charge; Animal or advanced skin model: induced inflammatory state with flare-triggered supervised dosing; evaluate epidermal thickness and biomarker change	Sweat variability and poor trigger specificity; mitigate by requiring multimodal confirmation, e.g., temp rise plus impedance or pH shift. Chemistry drift; mitigate with single-use calibrated sensor cartridges.	6–9 months	\\$20 \\$45

Concept	Complexity	Objective	Core components	Validation protocol	Main risks and mitigation	Timeline	Bud
Paper-battery or galvanic microneedle-iontophoresis scar patch	Medium-to-high	Develop a self-administered hybrid MN-iontophoresis patch for hypertrophic scar or difficult-barrier lesions	Solid or porous MN array, paper battery or galvanic cell, hydrogel depot, constant-current limiting, local temp monitoring, user-facing dose gauge; strongest precedents are the vaccine MN system and PBIMNP scar patch. ⁶¹	Mechanical tests: insertion force, fracture force, insertion depth; Ex vivo scar-mimicking tissue or dense collagen gel permeation; In vivo preclinical scar model if available; endpoints: uniformity of intratissue distribution, fibrosis markers, pain surrogate versus injection	MN fracture/regulatory burden; mitigate with mechanical proof testing and conservative needle geometry. Battery consistency; mitigate with current-limited design and lot QC.	8–12 months	\\$30 \\$70

Concept	Complexity	Objective	Core components	Validation protocol	Main risks and mitigation	Timeline	Bud
Stretch concept: reverse-iontophoresis sensing with supervisory closed-loop dosing	High	Create a research platform where reverse iontophoresis or microneedle sensing estimates tissue/metabolic state and supervises transdermal dosing	Reverse-iontophoresis glucose or drug sensor; impedance/temperature safety sensing; current driver; estimator software; optional hybrid delivery actuator rather than pure iontophoresis if payload is large; informed by reverse-iontophoresis reviews, glucose patents, and adjacent closed-loop diabetes patches. ⁶²	First validate sensing separately; second validate dose-by-charge model; third validate supervisory logic in silico and ex vivo; only then progress to a controlled preclinical study	Highest technical risk: poor correlation between non-invasive signal and therapeutic need. Regulatory risk is also highest. Mitigate by positioning as decision-support/supervisory research first, not autonomous treatment.	12–18 months	\\$60 \\$18

For most research groups, the highest-value sequence is **Concept one → Concept two or three → Concept four**, not Concept five first. That sequence preserves translational relevance while keeping physics, chemistry, and regulation aligned with what current papers already demonstrate. ⁶³

The Gantt below shows a realistic 9-month roadmap for a medium-complexity physiology-aware iontophoresis project. It is especially suitable for the dermatitis or pressure-ulcer concepts.



The most important immediate experimental steps are straightforward. First, pick a **single drug and single indication** and build the dose-by-charge and dose-by-impedance model before adding any physiological trigger. Second, validate **skin-interface stability** under sweat, motion, and long wear, because most elegant closed-loop logic collapses if the interface is unstable. Third, add only one actionable physiological input at first, preferably **skin impedance** or **temperature**, because those directly improve safety and controllability. Only after that should you add sweat chemistry, ISF biomarkers, or ML-based inference. ⁶⁴

Open questions remain because the initial brief left several clinically decisive items unspecified: the target drug, disease, patient age, body site, wear duration, caregiver supervision level, home-use versus clinic-use setting, and intended regulatory/geographic market. Those choices strongly affect acceptable current density, sensor choice, reservoir chemistry, adhesive strategy, software burden, and the likely U.S./EU pathway. A second major limitation is that this report is not a formal freedom-to-operate or clinical-practice recommendation; it is a research and development map grounded in recent literature, patent records, and regulator-facing guidance. ⁶⁵

A final high-confidence conclusion emerges from the literature: **the best near-term novel work is not “more current” or “more sensors,” but tighter integration of dose physics, skin-interface state, and simple safety-first control logic**. The systems most likely to succeed are those that combine physically grounded sensing such as impedance and temperature with a drug that genuinely suits iontophoretic transport, then add richer physiology only where it changes a real therapeutic decision. ⁶⁶

Selected primary and regulatory sources that should be at the top of any follow-on reading list are the 2021 review on integrating wearables with transdermal delivery, the 2023 review on wearable transdermal patches, the 2023 reverse-iontophoresis/flexible-electronics review, the 2022 pressure-ulcer sensing-treatment proof-of-concept, the 2023 Mg-battery patch, the 2023 vaccine microneedle-iontophoresis paper, the 2025 atopic-dermatitis patch, the 2025 paper-battery scar patch, the 2026 self-indicating eco-degradable patch, FDA’s transdermal/topical quality guidance, FDA’s iontophoresis-device classification pages, and EMA’s Article 117 and medicinal-product/device guidance. ⁶⁷

¹ ³ ⁵ ²⁴ ⁶⁶ ⁶⁷ <https://pubmed.ncbi.nlm.nih.gov/34634396/>
<https://pubmed.ncbi.nlm.nih.gov/34634396/>

² ⁶ ²³ ⁵⁷ <https://pubmed.ncbi.nlm.nih.gov/36690034/>
<https://pubmed.ncbi.nlm.nih.gov/36690034/>

⁴ <https://pubmed.ncbi.nlm.nih.gov/42198439/>
<https://pubmed.ncbi.nlm.nih.gov/42198439/>

⁷ ¹³ ¹⁵ ²¹ ³⁰ ³⁴ ⁵⁹ ⁶³ ⁶⁴ <https://www.sciencedirect.com/science/article/abs/pii/S1350453322001096>
<https://www.sciencedirect.com/science/article/abs/pii/S1350453322001096>

⁸ ²⁶ ²⁹ ³⁵ <https://www.nature.com/articles/s41467-023-35990-7>
<https://www.nature.com/articles/s41467-023-35990-7>

⁹ ³⁶ ⁶⁰ <https://pubmed.ncbi.nlm.nih.gov/39222610/>
<https://pubmed.ncbi.nlm.nih.gov/39222610/>

¹⁰ ²⁷ ³⁸ <https://www.nature.com/articles/s41467-021-20948-4>
<https://www.nature.com/articles/s41467-021-20948-4>

11 <https://pubmed.ncbi.nlm.nih.gov/39179551/>

<https://pubmed.ncbi.nlm.nih.gov/39179551/>

12 14 19 31 <https://www.nature.com/articles/s41551-021-00685-1>

<https://www.nature.com/articles/s41551-021-00685-1>

16 17 <https://pubmed.ncbi.nlm.nih.gov/39018979/>

<https://pubmed.ncbi.nlm.nih.gov/39018979/>

18 <https://pubmed.ncbi.nlm.nih.gov/38113695/>

<https://pubmed.ncbi.nlm.nih.gov/38113695/>

20 25 <https://pubmed.ncbi.nlm.nih.gov/37265867/>

<https://pubmed.ncbi.nlm.nih.gov/37265867/>

22 37 41 <https://www.nature.com/articles/s41528-026-00562-4>

<https://www.nature.com/articles/s41528-026-00562-4>

28 <https://pubmed.ncbi.nlm.nih.gov/29978889/>

<https://pubmed.ncbi.nlm.nih.gov/29978889/>

32 39 61 <https://www.nature.com/articles/s41378-023-00515-1>

<https://www.nature.com/articles/s41378-023-00515-1>

33 <https://www.nature.com/articles/s41378-024-00823-0>

<https://www.nature.com/articles/s41378-024-00823-0>

40 58 <https://pubmed.ncbi.nlm.nih.gov/23071949/>

<https://pubmed.ncbi.nlm.nih.gov/23071949/>

42 <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/transdermal-and-topical-delivery-systems-product-development-and-quality-considerations>

<https://www.fda.gov/regulatory-information/search-fda-guidance-documents/transdermal-and-topical-delivery-systems-product-development-and-quality-considerations>

43 46 <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpdc/classification.cfm?id=KTB>

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpdc/classification.cfm?id=KTB>

44 47 65 [How to Study and Market Your Device | FDA](#)

https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/how-study-and-market-your-device?utm_source=chatgpt.com

45 https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfstandards/detail.cfm?standard_identification_no=38829

https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfstandards/detail.cfm?standard_identification_no=38829

48 [Medicinal products used in combination with a medical device | European Medicines Agency \(EMA\)](#)

https://www.ema.europa.eu/en/medicinal-products-used-combination-medical-device?utm_source=chatgpt.com

49 56 <https://patents.google.com/patent/EP0702579B1/en>

<https://patents.google.com/patent/EP0702579B1/en>

50 <https://patents.google.com/patent/WO1998013100A1/en>

<https://patents.google.com/patent/WO1998013100A1/en>

51 <https://patents.google.com/patent/US9072883B2/en>

<https://patents.google.com/patent/US9072883B2/en>

⁵² <https://patents.google.com/patent/TW201825131A/en>
<https://patents.google.com/patent/TW201825131A/en>

⁵³ <https://patents.google.com/patent/CN116429854A/en>
<https://patents.google.com/patent/CN116429854A/en>

⁵⁴ <https://patents.google.com/patent/WO2024259501A1/en>
<https://patents.google.com/patent/WO2024259501A1/en>

⁵⁵ <https://patents.google.com/patent/US20190209828A1/en>
<https://patents.google.com/patent/US20190209828A1/en>

⁶² <https://pubmed.ncbi.nlm.nih.gov/36580817/>
<https://pubmed.ncbi.nlm.nih.gov/36580817/>