

Formulas used:

The reference resolution distance is calculated as follows:

$$res_{\theta,\varphi}^o = \frac{res_x^o * res_y^o * res_z^o}{\sqrt{(res_y^o * res_z^o * \cos\Phi * \sin\theta)^2 + (res_x^o * res_z^o * \sin\Phi * \sin\theta)^2 + (res_x^o * res_y^o * \cos\theta)^2}}$$

$$\Phi = \arccos \frac{x_B - x_A}{\sqrt{(x_B - x_A)^2 + (y_B - y_A)^2}}$$

$$\theta = \arccos \frac{z_B - z_A}{\sqrt{(x_B - x_A)^2 + (y_B - y_A)^2 + (z_B - z_A)^2}}$$

With

$$res_{x,y}^o = \frac{0.51 * \lambda_{ex}}{NA}$$

$$res_z^o = \frac{0.88 * \lambda_{ex}}{n - \sqrt{n^2 - NA^2}}$$

x_A, y_A, z_A and x_B, y_B, z_B are the bead coordinates in channel A and B respectively, NA: numerical aperture, λ_{em} : emission wavelength, n: refractive index of the lens immersion & mounting media.

Lateral ($res_{x,y}^o$) and axial (res_z^o) theoretical resolution values used for Spinning Disk microscopes are calculated as defined in Toomre and Pawley, Disk-Scanning Confocal Microscopy in Handbook Of Biological Confocal Microscopy 221–238 (Springer US, 2006).

Compliance with the Shannon-Nyquist criterion uses the formulas for widefield Shannon-Nyquist distances calculation:

$$\alpha = \arcsin\left(\frac{NA}{n}\right)$$

$$\Delta_{x,y} = \frac{\lambda_{ex}}{8 \cdot NA}$$

$$\Delta_z = \frac{\lambda_{ex}}{4 \cdot n \cdot (1 - \cos(\alpha))}$$