雌二醇校准品说明书

【产品名称】

通用名称：雌二醇校准品

英文名称：Access® Estradiol Calibrators

【包装规格】

校准品0（S0），4.0 mL/ 瓶； 校准品1（S1）–校准品5（S5），2.5 mL/ 瓶。

【预期用途】

本产品用于雌二醇项目测定的校准。

Access 雌二醇校准品用于校准Access Estradiol测定，以通过使用Access免疫测定系统定量测定人血清和血浆的雌二醇水平。S0可用于样本稀释。

【检验原理】

定量测定的校准是一个如同测试病人样本那样，对已知分析物浓度的样本（如，测定的校准品)进行测试以计算其反应值的过程。测得的反应值与已知分析物浓度之间的数学关系，可用于确立校准曲线。此数学关系，或校准曲线，用于将病人样本的RLU (光量子值) 测定值转换为特定的定量分析物浓度。

【主要组成成分】

校准品0（S0）:	人血清、< 0.1% 叠氮钠和 0.025% Cosmocil** CQ。含0 pg/mL(pmol/L) 雌二醇。
校准品1（S1）， 校准品2（S2）， 校准品3（S3）， 校准品4（S4）， 校准品5（S5）:	人血清中水平大约分别为106、570、1800、3100和4800 pg/mL(389、2092、6608、 11380和17621 pmol/L)的雌二醇（纯化合物)，含有 < 0.1% 叠氮钠和0.025% Cosmocil CQ。
校准卡：	1

溯源性

Access 雌二醇内的分析物，可溯源至ID/GC-MS（同位素稀释-气体色谱法/质谱分析）。溯源性的建立过程是按EN ISO 17511 进行的。

使用这一批次校准品中具有代表性的样本，建立指定的值，该值对于Access试剂的测定方法是特定的。使用其它方法确定的值可能会存在差异。若此差异存在，可能是不同方法之间的偏倚造成的。

【储存条件及有效期】

- 在 2~10℃条件下储存，有效期为 12 个月。
- 2~10℃冷藏保存且竖直放置。
- 2~10℃环境下可稳定保存 12 个月。

- 可能存在变质情况的表现为质控超出了范围。
- 生产日期和失效日期详见标签。

【适用仪器】

Access、Access2、Unicel Dxi600 Access和Unicel Dxi800 Access免疫分析系统

【检验方法】

1. 参考相应的系统操作手册和/ 或帮助系统，获取有关校准原理、设置校准品系列、校准品测试申请输入和评论校准数据的信息。
2. 使用前轻轻倒转，混匀成份。防止形成气泡。
3. Access Estradiol Calibrators提供的是6 种浓度 ：零和大约106、570、1800、3100和4800 pg/mL，是由合成雌二醇和人血清用重量法制备的。项目校准数据的有效期长达 14 天。校准品的测定S0和S1应做四份，S2-S5应做双份。
4. 参见校准卡以获取准确的浓度信息。

【检验方法的局限性】

试剂内如有微生物污染的迹象或显著浑浊，应弃去。

【注意事项】

- 供体外诊断使用。
- 经过测试发现用于制备试剂的人源性材料，对于乙型肝炎、丙型肝炎(HCV)、和人体免疫缺陷病毒(HIV-1 和HIV-2)，均显示为阴性或无反应。由于任何已知的测试方法均无法完全确保不存在感染性物质，因此处理试剂与病人样本时均视其具有传染疾病的可能。[16]
- 经过测试发现用于制备该产品的每种血清/ 血浆库，均对纤维蛋白原表现为阴性。
- 叠氮钠可能会与铅以及铜制的管件发生反应，形成具有高危爆炸性的金属叠氮物。在倾弃这种液体时，应以大量的水进行冲洗，防止叠氮物的集结[6]。
- 材料安全数据资料 (MSDS) 可按需索取。

【标识的解释】

制造商

EC REP
欧盟授权代表

生产日期

LOT
批次代码

REF
产品编号

温度极限

温度上限

温度下限

IVD
体外诊断医疗器械

警告

CONTROL -
阴性对照

CONTROL +
阳性对照

有效期

生物风险

含量足够测试<n>次

查阅使用说明

怕晒

【参考文献】

1 C A and Ashwood, Tietz Textbook of Clinical Chemistry, 2nd Edition. 1994. Edited by Burtis, E R Philadelphia, PA: W B Saunders Co., 1857-1863.
2 Yen, S S C , The Human Menstrual Cycle- Neuroendocrine regulation. In Reproductive Endocrinology. Edited by Yen, S S C and Jaffe, R B Philadelphia, PA: W B Saunders Co., 273-308, 1991.
3 Carr, B R , Disorders of the ovary and female reproductive tract. In Williams Textbook of Endocrinology, 8th edition. Edited by Wilson, J D and Foster, D W Philadelphia, PA: W B Saunders Co., 733-798, 1992.
4 Hall, J E Polycystic ovarian disease as a neuroendocrine disorder of the female reproductive axis. In Endocrinology and Metabolism Clinics of North America, Neuroendocrinology II. Edited by Veldhuis, J D Philadelphia, PA: W B Saunders Co., 75-92.
5 Reyes-Fuentes, A. and Veldhuis, J D 1993. Neuroendocrine physiology of the normal male gonadal axis. In Endocrinology and Metabolism Clinics of North America. Edited by Velduis, J D Philadelphia, PA: W.B. Saunders Co., 22:93-124.
6 DHHS (NIOSH) Publication No. 78-127, August 1976. Current Intelligence Bulletin 13 - Explosive Azide Hazard. Available http://www.cdc.gov/niosh.
7 Approved Guideline – Procedures for the Handling and Processing of Blood Specimens, H18-A3. 2004. Clinical and Laboratory Standards Institute.
8 Cembrowski GS, Carey RN. Laboratory quality management: QC QA. ASCP Press, Chicago, IL, 1989.
9 Kricka L. Interferences in immunoassays – still a threat. Clin Chem 2000; 46: 1037–1038.
10 Bjerner J, et al. Immunometric assay interference: incidence and prevention. Clin Chem 2002; 48: 613–621.
11 Williams, Textbook of Endocrinology, 8th edition. Edited by Wilson, J D and Foster, D W Philadelphia, PA: W B Saunders Co., 978 - 995, 1992.
12 Thienpont LM, Verhaeghe PG, Van Brussel KA, De Leenheer AP. Estradiol-17β quantified in serum by isotope dilution-gas chromatography-mass spectrometry: reversed phase C18 high performance liquid chromatography compared with immunoaffinity chromatography for sample pretreatment. Clin Chem 1988; 34: 2066-2069.
13 Thienpont LM, De Leenheer AP, Dirscherl CD. The certification of estradiol-17β in three lyophilized serum materials (CRM 576, CRM 577 and CRM 578). Commission of the European Communities, Community Bureau of Reference, BCR Information Report, EUR 17540 EN (1997).
14 Thienpont LM, Franzini C, Kratochvila J, Middle J, Ricos C, Siekmann L, et al. Analytical quality specifications for reference methods and operating specifications for networks of reference laboratories. Recommendations of the European EQA-Organizers Working Group B. Eur J Clin Chem Clin Biochem 1995; 33: 949-957.
15 Thienpont LM, Van Nieuwenhove B, Stockl D, Reinauer H, De Leenheer AP. Determination of reference method values by isotope dilution-gas chromatography/mass spectrometry: a five years' experience of two European reference laboratories. Eur J Clin Chem Clin Biochem 1996; 34: 853-860.
16 HHS Publication, 4th ed., April 1999. Biosafety in Microbiological and Biomedical

Laboratories. Available http://www.cdc.gov/od/ohs/biosfty/bmbl4/bmbl4toc.htm

【基本信息】

注册人/生产企业名称：Beckman Coulter, Inc. 贝克曼库尔特（美国）股份有限公司

住所：250 South Kraemer Blvd., Brea, CA 92821, USA

生产地址：1000 Lake Hazeltine Drive, Chaska, MN 55318-1084, USA

联系方式：001-800-742-2345

售后服务单位名称：贝克曼库尔特商贸（中国）有限公司

联系方式：021-38651000,4008855355

代理人名称：贝克曼库尔特商贸（中国）有限公司

住所：中国（上海）自由贸易试验区峨山路91弄98号3层01单元

联系方式：021-38651000,4008855355

【医疗器械注册证编号/产品技术要求编号】

国械注进20152403744

【说明书核准日期及修改日期】

2020-04-13

2020-11-27

2022-06-30