内因子抗体质控品说明书

【产品名称】

通用名称：内因子抗体质控品

英文名称：Intrinsic Factor Ab QC

【包装规格】

水平1（QC1）: 4.0mL/瓶×3瓶，水平2（QC2）: 4.0mL/瓶×3瓶。

【预期用途】

本产品用于内因子抗体检测时的质量控制。

【检验原理】

质控品可模拟病人样本的诸多特性，因此在监控Access Intrinsic Factor Ab免疫测定的系统性能时，显得十分重要。此外，它们是规范的实验室操作流程的一个组成部分。[14,23,24,25,26,27] 在对Access Intrinsic Factor Ab试剂进行检测时，需包括质控品以确认检测结果的准确性。若测试所使用的系统运行正常，那么所测试的值应在可接受的范围内。

【主要组成成分】

水平 1（QC1）：内因子抗体阴性的人血清，含<0.1%的叠氮钠和 0.5% ProClin 300，QC1 的值大约为 1.05AU/mL；

水平 2（QC2）：内因子抗体阳性的人血清，含小鼠蛋白、<0.1%的叠氮钠和 0.5% ProClin 300，QC2 的值大约为 2.00AU/mL；

质控卡（QC 值卡）。

【储存条件及有效期】

2～10℃保存，有效期 12 个月。

即用型。

2～10℃环境下保存时，稳定期可持续至标签上所指明的有效期。

首次使用后，把小瓶保存在 2～10℃环境下，稳定期可持续 56 天。

可能存在变质情况的表现为质控值超出标准。
参见 QC 值卡以获取均值和标准差（SD）。
生产日期和失效日期见包装标签。

【适用仪器】
适用于免疫测定系统Access、Access2、UniCel DxI 600 Access、UniCel DxI 800 Access。

【检验方法】
1．与测定病人样本一样，使用Access免疫检测系统来测定Access内因子抗体质控品内内因子抗体的浓度。
2．参考相应的系统操作手册和/或帮助系统，获取有关质控原理、设置质控品系列、质控品样本测试申请的输入和评论质控品数据的信息。
3． 使用前轻轻倒转，混匀成份。防止形成气泡。

【参考区间】
对于Access内因子抗体QC品值的指定，需选择整批内一些具有代表性的样本，并对它们进行检测以对均值做出可靠的预估。均值和标准差均列于QC值卡上。但由于技术、设备或试剂的差异，可能会导致与所列的值存在差距。因此，各实验室需确立个别适用的均值和标准差(SD)。

【检验方法的局限性】
与Access免疫测定系统及配套试剂使用。
试剂内如有微生物污染的迹象或显著浑浊，应弃去。

【注意事项】
• 供体外诊断使用。
• 按所描述的步骤进行操作，可使日常处理病人样本和血制品中所带来的风险降到最低。然而，无论这些产品的来源、处理方式或先前获得的证明，使用这些具有潜在传染性的物品均须按统一的注意事项和完善的临床实验室操作规范来进行。应使用合适的消毒剂来消除污染。遵循当地的操作规章和导则来保存并处理这些物品以及盛放这些物品的容器。
• 用于试剂准备的人源性物质已进行检测，并且为乙肝表面抗原（HbsAg），丙肝病毒抗体（HCV），以及人类免疫缺陷病毒（HIV－1以及HIV－2）阴性。由于任何已知

的检测方法均无法完全确保不存在感染性物质，因此处理试剂与病人样本时均视其具有传染疾病的可能[10]

• 在该产品制备中使用的每种血清/血浆库都已经进行检测，均为纤维蛋白原阴性。

• 叠氮钠可能会与铅以及铜制的管件发生反应，形成具有高危爆炸性的金属叠氮物。在倾弃这种液体时，应以大量的水进行冲洗，防止叠氮物的集结。[11]

• Xi.刺激物：0.5% ProClin 300

R 43：如接触皮肤可引起致敏作用。
S 28-37：接触到皮肤后应立即用大量肥皂和水冲洗。应戴适用的手套。

• 材料安全数据资料(MSDS) 可按需索取。

【参考文献】

1. Chanarin I. The Megaloblastic Anaemias. Second edition. London, Blackwell Scientific Publications, 1979.
2. Gueant JL. Autoantibodies in pernicious anemia type I patients recognize sequence 251-256 in human intrinsic factor. Proc Assoc Am Physicians 1997. 109 (5): 462-469.
3. Whittingham S, MacKay IR. Pernicious Anemia and Gastric Atrophy. In The Autoimmune Diseases. Academic Press; Orlando, Fla. 1985: 243-266.
4. Lindenbaum J. Status of laboratory testing in the diagnosis of megaloblastic anemia. Blood 1983. 61 (4):624-627.
5. Nimo RE, Carmel R, Increased sensitivity of detection of the blocking (type I) anti-intrinsic factor antibody.American Journal Clinical Pathology, 1987. 88 (6), 729-733.
6. Carmel R. Reassessment of the relative pervalences of antibodies to gastric parietal cell and to intrinsic factor inpatients with pernicious anaemia: influence of patient age and race. Clin. exp. Immunol. 1992. 89: 74-77.
7. Glass JB. Gastric Intrinsic Factor and Other Vitamin B12 Binders. Georg Thieme publishers Stuttgart, 1974:29-36.
8. Mardh S. Occurrence of autoantibodies against intrinsic factor, H, K-ATPase, and pepsinogen in atrophic gastritis and rheumatoid arthritis. Scand J of Gastroenterology, 1991. 26 (16):1089-96.
9. Chanarin I. The Megaloblastic Anaemias. Second edition. London, Blackwell Scientific Publications, 1979: 332.

10. HHS Publication, 4th ed., April 1999. Biosafety in Microbiological and Biomedical Laboratories. Available http://www.cdc.gov/od/ohs/biosfty/bmbl4/bmbl4toc.htm
11. DHHS (NIOSH) Publication No. 78-127, August 1976. Current Intelligence Bulletin 13 - Explosive Azide Hazard. Available http://www.cdc.gov/niosh.
12. Approved Guideline – Procedures for the Handling and Processing of Blood Specimens, H18-A3. 2004.Clinical and Laboratory Standards Institute.
13. Davidson RJL, Longitudinal study of circulating gastric antibodies in pernicious anaemia, J Clin Pathol, 1989. 42: 1092-1095.
14. Cembrowski GS, Carey RN. Laboratory quality management: QC QA. ASCP Press, Chicago, IL, 1989.
15. Kricka L. Interferences in immunoassays – still a threat. Clin Chem 2000; 46: 1037–1038.
16. Bjerner J, et al. Immunometric assay interference: incidence and prevention. Clin Chem 2002; 48: 613-621.
17. Chanarin I. The Megaloblastic Anaemias. Second edition. London, Blackwell Scientific Publications, 1979: 59.
18. Carmel R, The laboratory diagnosis of megalobalstic anemias. Wes J Med; 1978. 128 (4): 294-304.
19. Fairbanks VF. Tests for pernicious anemia: serum intrinsic factor blocking antibody. Proc Mayo Clin; 1983. 58:203-4.
20. Ungar B. Detection and titration of antibody to intrinsic factor by a modified charcoal method. Aus J Exper Biol Med Sci; 1967; 45: 317 -21.
21. Approved Guideline – Evaluation of precision performance of clinical chemistry devices, EP5-A. 1999.National Committee for Clinical Laboratory Standards.
22. Krouwer, JS, Rabinowitz, R. How to improve estimates of imprecision. Clinical Chemistry, 1984; 30: 290-292.
23. Broome HE, Cembrowski GS, Kahn SN, Martin PL, Patrick CA. Implementation and use of a manual multi-rule quality control procedure. Lab Med 1985; 16: 533-537.
24. Westgard JO, Barry PL, Hunt MR, Groth T. A multi-rule Shewhart chart for quality control in clinical chemistry.Clin Chem 1981; 27: 493-501.
25. Koch DD, Oryall JJ, Quam EF, Feldbruegger DH, et al. Selection of medically useful QC procedures for individual tests done in a multitest analytical system.

Clin Chem 1990; 36:230-233.

26. Mugan K, Carlson IH, Westgard JO. Planning QC procedures for immunoassays. J Clin Immunoassay 1994;17:216-222.

27. Approved Guideline – Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions, C24-A3. June 2006. Clinical and Laboratory Standards Institute.

【标识的解释】

制造商

欧盟授权代表

生产日期

批次代码

产品编号

温度极限

温度上限

温度下限

体外诊断医疗器械

警告

阴性对照

阳性对照

有效期

生物风险

含量足够测试<n>次

查阅使用说明

怕晒

【基本信息】

注册人/生产企业名称：Beckman Coulter, Inc. 贝克曼库尔特（美国）股份有限公司
住所：250 South Kraemer Blvd. Brea, CA 92821 USA
生产地址：1000 Lake Hazeltine Drive Chaska, MN 55318-1084 USA
联系方式：001-800-742-2345
售后服务单位名称：贝克曼库尔特商贸（中国）有限公司
联系方式：021-38651000
代理人名称：贝克曼库尔特商贸（中国）有限公司
住所：中国（上海）自由贸易试验区峨山路 91 弄 98 号 3 层 01 单元
联系方式：021-38651000

【医疗器械注册证编号/产品技术要求编号】

国械注进 20182402408

【说明书核准日期及修改日期】

2023-09-26